



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 285

Distribution No.: 156-A

Month/Year: April/2022

Instrument ID: Pentra ES60, Horiba 5 part diff hematoanalyzer, Serial no: 910PES15342

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 08-07-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	11.6	11.4	23	30.81	0.2690	-0.76	0.2	0.21	0.0130	-0.04
RBC x10 ⁶ /µl	1	6.03	5.99	12.02	12.23	0.0110	-0.66	0.04	0.06	0.0030	-0.39
Hb g/dl	1	13.1	13	26.1	25.9	0.0200	0.31	0.1	0.1	0.0070	0.00
HCT%	1	41.9	41.5	83.4	86.1	0.1480	-0.62	0.4	0.4	0.0240	0.00
MCV-fl	1	70	69	139	141.2	0.2110	-0.35	1	0.2	0.0200	2.16
MCH-Pg	1	21.9	21.6	43.5	42.4	0.0460	0.87	0.3	0.2	0.0100	0.67
MCHC-g/dl	1	31.4	31.3	62.7	59.8	0.1190	0.82	0.1	0.3	0.0200	-0.67
Plt. x10 ³ /µl	1	178	174	352	368	2.34	-0.23	4	9	0.53	-0.52
Retic %	2	12.8	12.6	25.4	15.85	0.27	1.13	0.2	0.4	0.02	-0.34

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=1 , Poly=47 L=48, E=1, Mono/Promono=4 , B1=0 P.M.=0, Mye=0, Meta=0, Other=Occasional hypersegmented neutrophils present	Lympho: 44-54, Poly: 40-49, Mono: 2-5, Eosino: 1-4, blast/Promyelo/Myelo/Meta: 0		
RBC Morphology	3	Many Macrocytes, few microcytes, moderate anisocytosis, few poikilocytes, some target cells, Occasional stomatocytes, moderate hypochromasia and few polychromasia.	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Microcytosis, Hypochromia; Mild: Poikilocytosis , Target cells , Sickle shaped cells , tear drop cells		
Diagnosis	3	Megaloblastic anaemia.	Diagnosis- Thalassemia/haemoglobinopathy		

Seema

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 156--A	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 ³ /µl	1	375	366	95.63	90.16	0.82	2.46	3.55	7.38
RBC x10 ⁶ /µl	1	375	375	84.27	87.2	7.73	4.53	8	8.27
Hb g/dl	1	375	375	89.87	86.67	4.27	5.07	5.86	8.26
HCT%	1	375	365	90.96	87.95	4.93	6.03	4.11	6.02
MCV-fl	1	375	366	90.96	92.6	5.75	1.92	3.29	5.48
MCH-Pg	1	375	366	89.07	89.07	7.1	4.64	3.83	6.29
MCHC-g/dl	1	375	366	91.8	89.34	4.37	3.01	3.83	7.65
Plt. x10 ³ /µl	1	375	366	93.44	90.98	4.1	3.55	2.46	5.47
ReticCount%	2	375	354	95.2	95.2	3.11	4.24	1.69	0.56
PS Assessment	3	375	345	Satisfactory :84.24%, Borderline Sat. :13.63%, Unsatisfactory :2.13%					

Comments:

1). Among Lab (EQA) : PS Diagnosis partially correct, remaining results acceptable

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----

*Z score of MCV + 2.16 lie ±2 to ±3
CAPA taken - Proper maintenance
done by Service
Engineer, Hoshiar*

*Parey
10/7/22*



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 285

Distribution No.: 155-A

Month/Year: January/2022

Instrument ID: Horiba Pentra ES 60 (Serial No-910PES15342)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 05-04-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	5.8	5.7	11.5	11.3	0.0260	0.27	0.1	0.1	0.0070	0.00
RBC x10 ⁶ /µl	1	4.35	4.32	8.67	8.94	0.0110	-0.79	0.03	0.04	0.0020	-0.22
Hb g/dl	1	13.7	13.6	27.3	27.3	0.0190	0.00	0.1	0.1	0.0060	0.00
HCT%	1	38.7	38.6	77.3	82.2	0.1300	-1.25	0.1	0.3	0.0210	-0.67
MCV-fl	1	89	89	178	185	0.2670	-0.82	0	0.3	0.0210	-0.67
MCH-Pg	1	31.5	31.4	62.9	60.7	0.0820	0.82	0.1	0.3	0.0180	-0.67
MCHC-g/dl	1	35.3	35.2	70.5	66.4	0.1090	1.11	0.1	0.3	0.0190	-0.67
Plt. x10 ³ /µl	1	230	228	458	434	0.95	0.88	2	5	0.32	-0.58
Retic %	2	2.6	2.5	5.1	4.2	0.09	0.30	0.1	0.2	0.01	-0.67

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT
DLC%	3	Nrbcs=2 , Poly=79 L=19, E=1, Mono/Promono=1 , B1=0 P.M.=0, Mye=0, Meta=0, Other=	Poly: 75-82 ,Lympho: 14-20, Mono: 2-4 , Eosino: 1-2, , Promyelo/Myelo/Meta/blast: 0
RBC Morphology	3	Reduce RBC count,Moderate anisocytosis,some macrocytes, few microcytes,few poikilocytes & tear drop cells,moderate hypochromia, and few polychromasia.	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Microcytosis, Hypochromia; Mild: Poikilocytosis
Diagnosis	3		Megaloblastic anemia

Raj
11/4/22

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 155--A	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 ³ /µl	1	390	389	86.63	82.78	5.4	9.51	7.97	7.71
RBC x10 ⁶ /µl	1	390	390	92.05	92.05	6.41	4.87	1.54	3.08
Hb g/dl	1	390	390	91.28	93.33	6.15	3.59	2.57	3.08
HCT%	1	390	390	93.85	91.28	4.36	4.1	1.79	4.62
MCV-fl	1	390	390	95.13	92.56	4.87	3.59	0	3.85
MCH-Pg	1	390	389	93.06	91.77	5.91	4.11	1.03	4.12
MCHC-g/dl	1	390	390	97.44	88.46	2.56	4.87	0	6.67
Plt. x10 ³ /µl	1	390	390	90.77	88.72	5.64	7.69	3.59	3.59
ReticCount%	2	390	390	87.18	75.9	4.36	11.03	8.46	13.07
PS Assessment	3	390	360	Satisfactory :96.42%, Borderline Sat. :3.07%, Unsatisfactory :0.51%					

Comments:

1). Among Lab (EQA) : PS Diagnosis not reported, remaining results acceptable

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Report authorized by,



Dr. Seema Tyagi (Prof.)
PT Co-ordinator: ISHTM-AIIMS-EQAP
Department of Hematology, AIIMS, New Delhi

-----End Of Report-----

11/14/22

V. Gupta

11/09/2022