



M31112041N

Quality Product and Service for Medical Laboratory & Biotech Instruments, Consumable & Reagents "An ISO 9001:2015 Certified Company"

SAGAR GALAXY, IInd Floor, LSC, Mayur Vihar - II, New Delhi -110091 Tel.: 011-41004431,

E-mail: priman@primanindia.com / email: admin@primanindia.com Website: www.primanindia.com

Date: 5th Apr 2022

CALIBRATION CERTIFICATE

This is to certify that Instrument model Fully Automated Biochemistry Analyzer Model BS-240 (Sr. No.YX-8A002065) Installed at Central Diagnostic Laboratory Address: Rajendra Memorial Institute of Medical Sciences, Patna, Bihar - 800007, has been calibrated by MINDRAY Biochemistry Calibrator MULTI SERA CALIBRATOR (Lot No. 150121003 EXP: 31-12-2022). During Calibration, Instrument connected through UPS.

Lamp Voltage: 12 V checked and found 11.98
Gain of all Filters are 48000- 60000.
As per the Manufacturer's Guideline the gain should be in range of (40,000- 60,000)

Next Calibration for the Instrument is due after six month from the date of this Certificate.

For Priman Instruments Pvt. Ltd.

Authorized Signatory

Date: 05/04/2022

Place: NEW DELHI