



RML – Quality Assurance Program (RML – QAP)



HEMATOLOGY ALL METHOD REPORT

Lab Code: 2564

Cycle: 11/2022
Round -4

Date: 01/08/2022

Complete Blood Count (CBC)

Parameters	No. of Participants	Group Mean	Standard deviation (SD)	Uncertainty of Assign Values	Range (± 2 SD)	Your Value	Standard Deviation Index(SDI)
Hb gm/dl	214	11.6	0.4	0.03	10.8-12.4	10.2	-3.5 →
WBC $\times 10^3/\mu\text{l}$.	212	10.5	2.4	0.21	5.6-15.3	11.1	0.3
RBC $\times 10^6/\mu\text{l}$.	214	4.0	0.1	0.01	3.74-4.30	3.64	-3.6 →
Hct%	213	34.9	2.4	0.21	30.2-39.6	31.7	-1.3
MCV fl.	214	86.8	4.2	0.36	78.5-95.2	87.2	0.1
MCH pg.	214	28.8	1.0	0.09	26.8-30.8	28.1	-0.7
MCHC gm/dl	214	33.1	2.0	0.17	29.1-37.2	32.2	-0.4
Platelet $\times 10^3/\mu\text{l}$.	214	269.3	22.6	1.93	224.0-314.5	246	-1.0

Interpretation of SDI:

SDI Value(+/-)	0 - 0.5	0.6 - 0.9	1.0 - 2.0	2.1 - 2.9	≥ 3
Interpretation	Excellent Performance	Good Performance	Acceptable Performance	Marginal Performance Need Improvement	Unacceptable Performance Needs Urgent action

Peripheral Blood Smear(PBS):

	Your Result	Consensus Result
DLC	-	L-66.0-84.7, P-11.8-24.7 ProL-3.0-13.3
Morphology	-	Δ Normocytic/ Normochromic (161/184) Δ Smudge cells (150/184) Δ Lymphocytosis (129/184) Δ Leukocytosis (101/184) Δ Atypical cells (30/184)
Diagnosis	-	Chronic Lymphocytic Leukemia (CLL), Chronic Lymphoproliferative Disorder (CLPD), Chronic Lymphoid Leukemia

Legends	(*) Excluded From Group Mean	(.) Not Reported	(#) Late Result Submission	(S) Reported in other Unit
---------	------------------------------	------------------	----------------------------	----------------------------

Chief Coordinator

Programme Director

Dr. Sanjay Mehrotra

Dr. Bandana Mehrotra

Checked By: *[Signature]*

Doc. No: ASS / FR / 06 / R.01 / Dt.: 05.01.2022

End of Report

Page 1 of 5

Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph : 4034100-130 (30 Lines), 4077180, 2788441 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence

