


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)



'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
 Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com CC-3427

Calibration Certificate

CERTIFICATE NO.	: 220902A/N-105	Page	: 1 of 1	
CALIBRATED ON.	: 02 / 09 / 2022	Ref. Std.	: DKD R-5-7	
CALIBRATION DUE.	: 01 / 09 / 2023	C.P. No.	: ACS/CP/11	
SRF No. & DATE	: 220902A/N Dt. 02.09.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000000497F Dt. 12.09.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 2 nd Floor, DAG No.238,243 Patta No,723,1166 Mauza Beltola Village: Odalbakra, Guwahati, Dist-Kamrup(M), Assam-781034			
ITEM CALIBRATED	: Digital Thermo Hygrometer			
	Make : HTC-1	Range	: (0 to 50)°C, (10-99)%	
	L. Count :	ID No.	: RCL/BIO/HGM-01	
	Location : Clinical Biochemistry	S. No.	: N.A	
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C} R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temp. Sensor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	15/07/2023
RESULTS	: ET Discipline			
* U. C. INSTT. READS	STANDARD READING	ERROR		
(°C)	25	25.3	- 0.3	
	28	28.4	- 0.4	
	32	32.5	- 0.5	
(%)	40	39	1	
	50	49	1	
	60	57	3	

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4°C, 1.5% at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmy statement not provided since not required.

Prepared by

....End of Calibration Certificate....

