



March 08, 2022

*Calibration Certificate for Urine Analyser  
Model: Laura (Sl. No. 0684-1-202106)*

*This is to certify that the Urine Analyser, Model: Laura installed at Redcliffe Lifetech PVT LTD, Guwahati has been calibrated. All parameters have been checked and found well within the limit.*

*This calibration is valid till March 07, 2023*

*For Transasia Bio-Medicals Ltd.*

*Taraknath Chakraborty,  
Dy. Regional Service Manager*



DEKAPHAN LAURA  
Seq.No: 0009  
URINORM N  
17.09.2022 12:45 PM

BLD NEG  
BIL NEG  
UBG NORM  
KET NEG  
GLU NEG  
PRO NEG  
PH 6  
NIT NEG  
LEU NEG  
SG 1.025



QC TEST: OK

DEKAPHAN LAURA  
Seq.No: 0010  
URINORM P  
17.09.2022 12:48 PM

\*BLD 50 Ery/u1  
\*BIL 3 mg/dl  
\*UBG 1 mg/dl  
\*KET 52 mg/dl  
\*GLU 250 mg/dl  
\*PRO 500 mg/dl  
PH 9  
\*NIT POS  
\*LEU 75 Leu/u1  
SG 1.000



QC TEST: OK