


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 220902A/N-106	Page	: 1 of 1
CALIBRATED ON.	: 03 / 09 / 2022	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 02 / 09 / 2023	C.P. No.	: ACS/CP/10
SRF No. & DATE	: 220902A/N Dt. 02.09.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342722000000498F Dt. 12.09.2022	Cal. At	: Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
 2nd Floor, DAG No.238,243 Patta No,723,1166 Mauza Beltola
 Village: Odalbakra, Guwahati, Dist-Kamrup(M), Assam-781034

ITEM CALIBRATED	: Centrifuge
Make	: Remi
Range	: (0 to 5250) RPM
L. Count	: ID No. : RCL/CP/Cent-01
Location	: Clinical Pathology
	: S. No. : ZHHN-33098

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ TM5010EK-039	Godrej	M-210914-16-1	14.09.2022

RESULTS :			
Mechanical Discipline			
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.68	- 1.68	
2000	2002.20	- 2.20	
3000	3003.48	- 3.48	
4000	4005.36	- 5.36	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

