


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 220902A/N-107	Page	: 1 of 1
CALIBRATED ON.	: 03 / 09 / 2022	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 02 / 09 / 2023	C.P. No.	: ACS/CP/10
SRF No. & DATE	: 220902A/N Dt. 02.09.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342722000000499F Dt. 12.09.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 2 nd Floor, DAG No.238,243 Patta No,723,1166 Mauza Beltola Village: Odalbakra, Guwahati, Dist-Kamrup(M), Assam-781034		

ITEM CALIBRATED	: Centrifuge			
	Make	: Remi	Range	: (0 to 5250) RPM
	L. Count	:	ID No.	: RCL/BIO/Cent-02
	Location	: Clinical Biochemistry	S. No.	: ZHFN-22117

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ TM5010EK-039	Godrej	M-210914-16-1	14.09.2022

RESULTS	: Mechanical Discipline		
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.76	- 1.76	
2000	2002.32	- 2.32	
3000	3003.54	- 3.54	
4000	4005.48	- 5.48	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

