



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2022



PC-1024

Lab Name	SHREERAKSHA LABS PRIVATE LIMITED	Lab No	11265
Constituent Group	HbA1c	Date of Result Entered :	08/08/2022
PT item	Lyophilized Whole blood	Date of Report Published :	05/09/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	110	5.69	9.80	0.56	5.6 %	-0.16	0.11

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

SHREERAKSHA LABS PRIVATE LIMITED
 NO:46, 17th CROSS BETWEEN MARGOSA ROAD & SAMPIGE ROAD, MALLESHWARAM
 BANGALORE
 KARNATAKA560055

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****