

Amogh Diagnostic Solutions Pvt. Ltd.


The Scientific Solution Providers



CALIBRATION CERTIFICATE								
Customer Name	Department of Haematology-Laboratory Medicine No:#46,17th cross ,Malleshwaram,Bangalore-560055							
Department								
Address	No:#46,17th cross ,Ma	Heshwaram, Bangalore-300033						
		Date of Calibration	27.07.2022					
Instrument	Mindray BC-5130 ANALYZER	Next Calibration Date	26.07.2023					
Serial No	TR-21006640	Calibrator Lot Number	PLUS0722					
		Calibrator Expiry date	05.08.2022					

Calibration Values	<u>Parameter</u>	<u>Target</u>	<u>Range</u>	Mean of Measured Values	<u>New</u> <u>Factor</u>	Old Factor
	WBC x 10 ⁹ /L RBC x10 ¹² /L	9.04	8.84-9.24	8.84	102.23	100.00
		4.57	4.49-4.65	4.39	104.08	100.00
	HGB g/L	13.5	13.3-13.7	13.3	101.45	100.00
	MCV fl	89.4	87.4-91.4	90.3	99.03	100.00
	PLT x 10 ⁹ /L	254	242-266	244	104.04	100.00

This is to certify that the Mindray BC 5130 Hematology Analayzer Sr.No: TR-21006640system, installed at Shree Raksha Labs Pvt Ltd , Bangalore has been calibrated using the standard reference material SC CAL PLUS-0522 Calibrator. Traceablity of SC CAL PLUS Calibrator material from the manufacturer is attached along with this certificate.



Application Specialist-IVD Mindray Medical India Pvt.Ltd Date:27.07.2022