



Certificate No.: TC/22/9539-03

Page 1 of 2

ULR-CC234422000021567F

CALIBRATION CERTIFICATE

- Name & Address of the Customer** : M/s. Sheeraksha Labs Private Limited.,
#46, 17th Cross, Margosa Road, Malleswaram,
Bangalore-560003
- Customer Reference**
Service Request No. : 9539
Date of Receipt : 30 July 2022
Date of Issue : 01 August 2022
- Details of Device Under Calibration (DUC)**
Nomenclature : Freezer
Make : Blue Star
Model : CHF 220
Serial No. : BKG 39089 L07
Id No. : ---
Location : 3rd Floor
DUC Condition : Satisfactory
SOP Number : TC/SOP/108
No. of Pages : 2
Calibration Date : 30 July 2022
Calibration Due : 29 July 2023
Calibration Done At : ONSITE
- Environmental Conditions**
Temperature : 25 ± 4 °C Humidity : 30-75 % Rh
- Standards Used for Calibration (STD)**

Sl. No.	Nomenclature	Serial No.	Traceable to	Certification No.	Validity
1	RTD Sensors with Paperless Recorder	031 to 048 & VR-020651	TestCAL	TC/21/106-04	25 August 2022

6. Note

- The Standard and equipments used for the calibration are Traceable to National / International Standards
- Any error in this Certificate must be brought to the knowledge of TestCAL within 30 days of issue.
- Results reported are valid at the time of & under the stated conditions of measurement

Calibrated by

Vignesh GS

Calibration Engineer



Authorized by

Srikanth G

Chief Operating officer



Certificate No.: TC/22/9539-03

Page 2 of 2

ULR-CC234422000021567F

7. Results:

Range : -8 to -20 °C


Resolution : 0.1 °C

Sl. No.	DUC Set in °C	DUC Reading in °C	Standard Sensors Reading in °C					Temp. Uniformity with Respect to Set Point in °C	
			1	2	3	4	5	+ Side	- Side
1	-20.0	-20.0	-20.5	-20.3	-20.5	-20.3	-20.2	0.0	0.5

Note: Stability of the freezer is \pm 0.50
Measurement Uncertainty \pm 0.70

8. Pictorial Representation of Sensors Placement in

Freezer



9. Conclusion Remarks:


1. Measurement Uncertainty reported is at 95.45% confidence level with $k = 2$.
2. Calibration points were selected as per customer specifications.

Calibrated by


Vignesh GS
Calibration Engineer



Authorized by


Srikanth G
Chief Operating officer