

Certificate No.: TC/22/9539-02
ULR-CC234422000021566F

Page 1 of 3

CALIBRATION CERTIFICATE

- Name & Address of the Customer** : M/s. Sheeraksha Labs Private Limited.,
#46, 17th Cross, Margosa Road, Malleswaram,
Bangalore-560003
- Customer Reference**
Service Request No. : 9539
Date of Receipt : 30 July 2022
Date of Issue : 01 August 2022
- Details of Device Under Calibration (DUC)**
Nomenclature : Refrigerator
Make : Elampro
Model : ECG-400
Serial No. : J27401001190009
ID No. : ----
Location : Sheeraksha Labs Private Limited 3rd Floor
DUC Condition : Satisfactory
SOP Number : TC/SOP/109
No. of Pages : 3
Calibration Date : 30 July 2022
Calibration Due : 29 July 2023
Calibration Done At : ONSITE
- Environmental Conditions**
Temperature : 25 ± 4 °C Humidity : 30-75 % Rh
- Standards Used for Calibration (STD)**

Sl. No.	Nomenclature	Serial No.	Traceable to	Certification No.	Validity
1	RTD Sensors with Paperless Recorder	031 to 048 & VR-020651	TestCAL	TC/21/106-04	25 August 2022

6. **Note**

- The Standard and equipments used for the calibration are Traceable to National / International Standards
- Any error in this Certificate must be brought to the knowledge of TestCal within 30 days of issue.
- Results reported are valid at the time of & under the stated conditions of measurement

Calibrated by

Vignesh GS
Calibration Engineer

Authorized by

Srikanth G
Chief Operating officer



Certificate No.: TC/22/9539-02

Page 2 of 3

ULR-CC234422000021566F

7. Results:


Range: 1 to 10 °C

Working Range : 2 to 8 °C

Date	Time	Freezer Std Sensors Reading in (°C)		Refrigerator Std Sensors Reading in (°C)		
		1	2	3	4	5
30-07-2022	11:45:00 AM	5.9	5.9	6.0	6.2	6.2
30-07-2022	11:46:00 AM	5.9	5.9	6.0	6.2	6.2
30-07-2022	11:47:00 AM	5.9	5.9	6.0	6.2	6.2
30-07-2022	11:48:00 AM	5.9	5.9	5.9	6.2	6.2
30-07-2022	11:49:00 AM	5.9	5.9	5.9	6.2	6.2
30-07-2022	11:50:00 AM	5.9	5.9	5.9	6.2	6.2
30-07-2022	11:51:00 AM	5.9	5.9	5.9	6.2	6.2
30-07-2022	11:52:00 AM	5.9	5.9	5.9	6.2	6.2
30-07-2022	11:53:00 AM	5.9	5.9	5.9	6.2	6.2
30-07-2022	11:54:00 AM	5.8	5.9	5.9	6.1	6.2
30-07-2022	11:55:00 AM	5.8	5.8	5.9	6.1	6.1
30-07-2022	11:56:00 AM	5.8	5.8	6.0	6.1	6.1
30-07-2022	11:57:00 AM	5.8	5.8	6.0	6.1	6.1
30-07-2022	11:58:00 AM	5.8	5.8	6.0	6.1	6.1
30-07-2022	11:59:00 AM	5.8	5.8	6.0	6.1	6.1
30-07-2022	12:00:00 PM	5.8	5.8	6.0	6.1	6.1
30-07-2022	12:01:00 PM	5.8	5.8	6.0	6.1	6.1
30-07-2022	12:02:00 PM	5.8	5.8	6.0	6.1	6.1
30-07-2022	12:03:00 PM	5.8	5.8	5.9	6.1	6.1
30-07-2022	12:04:00 PM	5.8	5.8	6.0	6.1	6.1
30-07-2022	12:05:00 PM	5.9	5.8	6.0	6.2	6.1
30-07-2022	12:06:00 PM	5.9	5.8	6.0	6.2	6.1
30-07-2022	12:07:00 PM	5.9	5.8	5.9	6.2	6.1
30-07-2022	12:08:00 PM	5.9	5.7	6.1	6.2	6.0
30-07-2022	12:09:00 PM	5.9	5.7	6.1	6.2	6.0
30-07-2022	12:10:00 PM	5.9	5.7	6.2	6.2	6.0
30-07-2022	12:11:00 PM	5.9	5.8	6.2	6.2	6.1
30-07-2022	12:12:00 PM	5.9	5.8	6.2	6.2	6.1
30-07-2022	12:13:00 PM	5.9	5.8	6.2	6.2	6.1
30-07-2022	12:14:00 PM	5.9	5.8	6.2	6.2	6.1
30-07-2022	12:15:00 PM	5.9	5.8	6.3	6.2	6.1
30-07-2022	12:16:00 PM	5.9	5.9	6.3	6.2	6.2
30-07-2022	12:17:00 PM	5.9	5.9	6.3	6.2	6.2

Average:	5.9	5.8	6.0	6.2	6.1
Max Temp:	5.9	5.9	6.3	6.2	6.2
Min Temp:	5.8	5.7	5.9	6.1	6.0

Calibrated by


Vignesh GS
Calibration Engineer



Authorized by



Srikanth G
Chief Operating officer



Certificate No.: TC/22/9539-02
ULR-CC234422000021566F

Page 3 of 3

8. Pictorial Representation of Sensors Placement in Refrigerator




Note: Stability of a Refrigerator is found to be ± 0.4 °C
Expanded Uncertainty is ± 1.2 °C

9. Conclusion Remarks:


- 1 Measurement Uncertainty reported is at 95.45% confidence level with $k = 2$.
- 2 Calibration points were selected as per customer specifications.

Calibrated by


Vignesh GS
Calibration Engineer



Authorized by


Srikanth G
Chief Operating officer