



Hemoglobin (Hb1Ac)

ALL METHOD REPORT

Cycle - 11/2022

Date: 07/09/2022

Lab Code: 2544

Round No - 3

Parameters	Units	No.of. Parti- cipants	Group Mean	Standard deviation (SD)	Uncertainty of Assign Values	Range (± 2 SD)	Your Value	Standard Deviation Index (SDI)
Hemoglobin (Hb A1C)	%	50	9.1	1.7	0.30	5.8-12.5	10.9	1.1

Interpretation of SDI:

SDI Value(+/-)	0 - 0.5	0.6 - 0.9	1.0 - 2.0	2.1 - 2.9	≥ 3
Interpretation	Excellent Performance	Good Performance	Acceptable Performance	Marginal Performance Needs Improvement	Unacceptable Performance Needs Urgent action

Legends	(*) Excluded From Group	[.] Not Reported	(#)Late Result	(\$)Reported in
---------	-------------------------	------------------	----------------	------------------


Chief Coordinator


Dr.Sanjay Mehrotra

Checked By

Prepared by:PS

Programme Director


Dr.Bandana Mehrotra

Page 1 of 1

