

ADVANCE CALIBRATION SERVICES


(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



CC-3427



NABL ACCREDITED LAB
CC-3427

Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220909A/N-116 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 10 / 09 / 2022 | Ref. Std. | : CG-11 | |
| CALIBRATION DUE. | : 09 / 09 / 2023 | C.P. No. | : ACS/CP/07 | |
| SRF No. & DATE | : 220909A/N Dt. 09.09.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000523F Dt. 14.09.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. Goyal Nursing Home and Hospital, Opposite Hanuman Mandir Sector-8, Faridabad | | | |
| ITEM CALIBRATED | : Digital Temperature Indicator | | | |
| | Make : RoHs | Range | : (-50 to 110) °C | |
| | L. Count : | ID No. | : RCL/HEM/D.Temp-04 | |
| | Location : Hematology | S. No. | : N.A | |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2021/0784 | 11/11/2022 |
| RESULTS | : ET Discipline | | | |
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) | | |
| 2.02 | 2 | 0.02 | | |
| 4.04 | 4 | 0.04 | | |
| 6.06 | 6 | 0.06 | | |
| 8.08 | 8 | 0.08 | | |
| (average of readings) | | | | |

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

