



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



CC-3427



Calibration Certificate

CERTIFICATE NO.	: 220909A/N-117	Page	: 1 of 1	
CALIBRATED ON.	: 10 / 09 / 2022	Ref. Std.	: CG-11	
CALIBRATION DUE.	: 09 / 09 / 2023	C.P. No.	: ACS/CP/07	
SRF No. & DATE	: 220909A/N Dt. 09.09.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000000524F Dt. 14.09.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Goyal Nursing Home and Hospital, Opposite Hanuman Mandir Sector-8, Faridabad			
ITEM CALIBRATED	: Digital Temperature Indicator			
	Make : RoHS	Range	: (-50 to 110) °C	
	L. Count :	ID No.	: RCL/SR/D.Temp-05	
	Location : Sample Receiving Room	S. No.	: N.A	
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784	11/11/2022
RESULTS	: ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
2.04	2	0.04		
4.08	4	0.08		
6.10	6	0.10		
8.12	8	0.12		
	(average of readings)			

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is $\pm 1.00^\circ\text{C}$ at coverage Factor $k=2$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

...End of Calibration Certificate....

