



Certificate of Calibration


Customer Name : Redcliffe Lifetech Pvt.Ltd Faridabad.
Model : Biochemistry Analyzer, Erba EM 200
Serial No. 2002949
Calibration Date : 14-04-2022
Next Calibration Due Date : 13-04-2023

With reference to the Semi Automated Biochemistry Analyser, Model :ERBA EM 200 bearing Sr. No. 2002949 at your lab, on inspection of the instrument it is observed that the results are well within the range and instrument is working fine.

Instrument is properly calibrated.

Thanking you,
For Transasia Bio-Medicals Ltd.

Devender singh
Sr. Service Engineer



Result Reprint

Report Type : Controls

Consumable: BIORAD L2

Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
89702	BIORAD L2	SGPTD	91.3 U/L		26708	20-Sep-2022 05:09:41	94,700	9,500	66.2 - 123.2
89702	BIORAD L2	SGOTD	175.0 U/L		26712	20-Sep-2022 05:10:53	187,000	18,000	133 - 241
89702	BIORAD L2	ALPU	562 U/L		26716	20-Sep-2022 05:12:05	517,000	51,000	364 - 670
89702	BIORAD L2	GGT	204.1 U/L	+1SD	26720	20-Sep-2022 05:13:17	176,000	18,000	122 - 230
89702	BIORAD L2	PHOS	6.63 mg/dL	-1SD	26724	20-Sep-2022 05:14:29	7,500	0,600	5.7 - 9.3
89702	BIORAD L2	BIDD	1.11 mg/dL	-1SD	26728	20-Sep-2022 05:15:42	1,280	0,160	0.8 - 1.76
89702	BIORAD L2	BIT	4.38 mg/dL		26732	20-Sep-2022 05:16:54	4,260	0,540	2.64 - 5.88
89702	BIORAD L2	UREA	104.0 mg/dL		26736	20-Sep-2022 05:18:06	100,000	7,500	77.5 - 122.5
89702	BIORAD L2	CRENZ	4.98 mg/dL		26740	20-Sep-2022 05:19:18	5,200	0,450	3.85 - 6.55
89702	BIORAD L2	GLU	280.1 mg/dL		26744	20-Sep-2022 05:20:30	291,000	22,500	223.5 - 358.5
89702	BIORAD L2	TRIG	82.7 mg/dL		26748	20-Sep-2022 05:21:43	85,700	6,450	66.35 - 105.05
89702	BIORAD L2	CHOL	98 mg/dL		26752	20-Sep-2022 05:22:55	101,000	7,750	77.75 - 124.25
89702	BIORAD L2	HDLC	22.5 mg/dL		26754	20-Sep-2022 05:23:49	23,000	2,250	16.25 - 29.75
89702	BIORAD L2	LDL	63.5 mg/dL		26756	20-Sep-2022 05:24:25	66,700	6,600	46.9 - 86.5
89702	BIORAD L2	UA	10.8 mg/dL		26760	20-Sep-2022 05:25:37	10,400	0,750	8.15 - 12.65
89702	BIORAD L2	CA	13.0 mg/dL		26764	20-Sep-2022 05:26:49	13,000	1,050	9.85 - 16.15
89702	BIORAD L2	ALBD	2.44 g/dL	-1SD	26768	20-Sep-2022 05:28:02	2,700	0,210	2.07 - 3.33

Print Date : 20-Sep-2022

Result Reprint

Report Type : Controls

Consumable: BIORAD L2

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
18	89702	BIORAD L2	PRO	4.17 g/dL	-1SD	26772	20-Sep-2022 05:29:14	4.600	0.300	3.7 - 5.5
19	89702	BIORAD L2	FE	60.1 µg/dL		26774	20-Sep-2022 05:30:08	60.200	6.000	42.2 - 78.2



Result Reprint

Consumable: BIORAD L1

Report Type : Controls

Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
16491	BIORAD L1	SGPTD	25.4 U/L		26707	20-Sep-2022 05:09:23	27.400	2.700	19.3 - 35.5
16491	BIORAD L1	SGOTD	36.1 U/L		26711	20-Sep-2022 05:10:35	37.100	3.750	25.85 - 48.35
16491	BIORAD L1	ALPU	124 U/L		26715	20-Sep-2022 05:11:47	115.000	12.000	79 - 151
16491	BIORAD L1	GGT	76.9 U/L	+1SD	26719	20-Sep-2022 05:12:59	66.300	6.600	46.5 - 86.1
16491	BIORAD L1	PHOS	3.44 mg/dL	-1SD	26723	20-Sep-2022 05:14:11	3.870	0.280	3.03 - 4.71
16491	BIORAD L1	BIDD	0.40 mg/dL		26727	20-Sep-2022 05:15:24	0.390	0.050	0.24 - 0.54
16491	BIORAD L1	BIT	1.02 mg/dL		26731	20-Sep-2022 05:16:36	1.010	0.120	0.65 - 1.37
16491	BIORAD L1	UREA	31.5 mg/dL		26735	20-Sep-2022 05:17:48	33.800	2.550	26.15 - 41.45
16491	BIORAD L1	CRENZ	1.65 mg/dL		26739	20-Sep-2022 05:19:00	1.660	0.160	1.18 - 2.14
16491	BIORAD L1	GLU	81.9 mg/dL		26743	20-Sep-2022 05:20:12	85.400	6.450	66.05 - 104.75
16491	BIORAD L1	TRIG	175.5 mg/dL		26747	20-Sep-2022 05:21:25	182.000	13.500	141.5 - 222.5
16491	BIORAD L1	CHOL	236 mg/dL		26751	20-Sep-2022 05:22:37	251.000	19.500	192.5 - 309.5
16491	BIORAD L1	HDLC	60.3 mg/dL	-1SD	26753	20-Sep-2022 05:23:31	67.200	6.750	46.95 - 87.45
16491	BIORAD L1	LDL	144.2 mg/dL		26755	20-Sep-2022 05:24:07	149.000	15.000	104 - 194
16491	BIORAD L1	UA	5.5 mg/dL	-1SD	26759	20-Sep-2022 05:25:19	6.440	0.480	5 - 7.88
16491	BIORAD L1	CA	9.6 mg/dL		26763	20-Sep-2022 05:26:31	9.210	0.690	7.14 - 11.28
16491	BIORAD L1	ALBD	3.66 g/dL		26767	20-Sep-2022 05:27:44	3.960	0.300	3.06 - 4.86

Result Reprint

Report Type : Controls

Consumable: BIORAD L1

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
18	26491	BIORAD L1	PRO	6.21 g/dL		26771	20-Sep-2022 05:28:56	6.510	0.500	5.01 - 8.01
19	26491	BIORAD L1	FE	265.9 µg/dL	+1SD	26773	20-Sep-2022 05:29:50	236.000	24.000	164 - 308



Result Reprint

Report Type : Controls

Consumable: HBA1C L

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
1	S022219	HBA1C L	HA1cD	5.3 %		26790	20-Sep-2022 06:23:35	5.500	0.400	4.3 - 6.7



Result Reprint

Report Type : Controls

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
1	S032224	HBA1C H	HA1cD	13.3 %	F	28805	20-Sep-2022 07:09:25	13.400	1.000	10.4 - 16.4

