

CALIBRATION CERTIFICATE			
Customer Name	<i>Shree Raksha Labs pvt ltd-Bangalore</i>		
Department	Department of Haematology-Laboratory Medicine		
Address	No:#46,17th cross ,Malleshwaram,Bangalore-560055		
		Date of Calibration	27.07.2022
Instrument	<i>Mindray BC-5130 ANALYZER</i>	Next Calibration Date	26.07.2023
Serial No	TR-21006640	Calibrator Lot Number	PLUS0722
		Calibrator Expiry date	05.08.2022

<u>Calibration Values</u>	<u>Parameter</u>	<u>Target</u>	<u>Range</u>	<u>Mean of Measured Values</u>	<u>New Factor</u>	<u>Old Factor</u>
		WBC x 10 ⁹ /L	9.04	8.84-9.24	8.84	102.23
	RBC x10 ¹² /L	4.57	4.49-4.65	4.39	104.08	100.00
	HGB g/L	13.5	13.3-13.7	13.3	101.45	100.00
	MCV fl	89.4	87.4-91.4	90.3	99.03	100.00
	PLT x 10 ⁹ /L	254	242-266	244	104.04	100.00

This is to certify that the Mindray BC 5130 Hematology Analyzer Sr.No: TR-21006640system, installed at Shree Raksha Labs Pvt Ltd , Bangalore has been calibrated using the standard reference material SC CAL PLUS-0522 Calibrator. Traceability of SC CAL PLUS Calibrator material from the manufacturer is attached along with this certificate.



Application Specialist-IVD Date:27.07.2022
Mindray Medical India Pvt.Ltd