

mindray


healthcare within reach

CALIBRATION CERTIFICATE

Customer Name	Shree Raksha Labs pvt ltd-Bangalore		
Department	Department of Haematology-Laboratory Medicine		
Address	No:#46,17th cross ,Malleshwaram,Bangalore-560055		
		Date of Calibration	27.07.2022
Instrument	<i>Mindray BC-5130 ANALYZER</i>	Next Calibration Date	26.07.2023
Serial No	TR-21006640	Calibrator Lot Number	PLUS0722
		Calibrator Expiry date	05.08.2022

Calibration Values	Parameter	Target	Range	Mean of Measured Values	New Factor	Old Factor
		WBC x 10 ⁹ /L	9.04	8.84-9.24	8.84	102.23
	RBC x10 ¹² /L	4.57	4.49-4.65	4.39	104.08	100.00
	HGB g/L	13.5	13.3-13.7	13.3	101.45	100.00
	MCV fl	89.4	87.4-91.4	90.3	99.03	100.00
	PLT x 10 ⁹ /L	254	242-266	244	104.04	100.00

This is to certify that the Mindray BC 5130 Hematology Analyzer Sr.No: TR-21006640system, installed at Shree Raksha Labs Pvt Ltd , Bangalore has been calibrated using the standard reference material SC CAL PLUS-0522 Calibrator. Traceability of SC CAL PLUS Calibrator material from the manufacturer is attached along with this certificate.



Application Specialist-IVD
Mindray Medical India Pvt.Ltd

Date:27.07.2022

mindray

healthcare within reach

CALIBRATION CERTIFICATE

Customer Name	Shree Raksha Labs pvt ltd-Bangalore		
Department	Department of Haematology-Laboratory Medicine		
Address	No:#46,17th cross ,Malleshwaram,Bangalore-560055		
		Date of Calibration	27.07.2022
Instrument	<i>Mindray BC-5130 ANALYZER</i>	Next Calibration Date	26.07.2023
Serial No	TR-21006640	Calibrator Lot Number	PLUS0722
		Calibrator Expiry date	05.08.2022

<u>Calibration Values</u>	<u>Parameter</u>	<u>Target</u>	<u>Range</u>	<u>Mean of Measured Values</u>	<u>New Factor</u>	<u>Old Factor</u>
	WBC x 10 ⁹ /L	9.04	8.84-9.24	8.84	102.23	100.00
	RBC x10 ¹² /L	4.57	4.49-4.65	4.39	104.08	100.00
	HGB g/L	13.5	13.3-13.7	13.3	101.45	100.00
	MCV fl	89.4	87.4-91.4	90.3	99.03	100.00
	PLT x 10 ⁹ /L	254	242-266	244	104.04	100.00

This is to certify that the Mindray BC 5130 Hematology Analyzer Sr.No: TR-21006640system, installed at Shree Raksha Labs Pvt Ltd , Bangalore has been calibrated using the standard reference material SC CAL PLUS-0522 Calibrator. Traceability of SC CAL PLUS Calibrator material from the manufacturer is attached along with this certificate.

Application Specialist-IVD
Mindray Medical India Pvt.Ltd

Date:27.07.2022

PERFORMANCE QUALIFICATION

BC 5130

AUTOMATED HAEMATOLOGY

ANALYSER

INSTALLED AT SHREE RAKSHA LABS PVT LTD
No:#46,17th Cross, Malleswaram, Bangalore, Karnataka- 560055




Type of Instrument:

Module.	Serial No
BC-5130	TR-21006640

PERFORMANCE QUALIFICATION VERIFICATION REPORT

This is to certify that the following checks mentioned in the performance verification protocol have been performed and found to be satisfactory.

PERFORMANCE VERIFICATION		
1	BACKGROUND	PASS
2	PRECISION (REPRODUCIBILITY) CHECK	PASS
3	CARRY OVER	PASS
4	CALIBRATION	PASS



Representative Signature

Date: 14.09.2022

Customer Signature

Date:

PERFORMANCE QUALIFICATION

Performance Specifications

The following performance specifications apply to systems that have been installed and maintained according to the guidelines in this manual and are operated with the recommended reagents and supplies. Specifications listed apply to all modes and test selections. System performance is expected to meet or exceed the specifications listed.

Checking before Calibration

Background

Background concentrations represent apparent sample-related constituents that actually originate from blood-free reagents and/or electronic “noise.” The background concentrations are used to confirm the System’s baseline performance, where no actual sample is aspirated. The following table lists acceptable background concentration limits that must be met before using the instrument.

Before calibration, make sure the background (blank count) results, repeatability results and carryover results are all within the specified ranges.

Background/blank count requirements for blood samples

Parameters	Acceptable range
WBC	$\leq 0.2 \times 10^9 / L$
RBC	$\leq 0.02 \times 10^{12} / L$
HGB	$\leq 1 \text{ g} / L$
PLT	$\leq 10 \times 10^9 / L$

Precision (Reproducibility)

Precision refers to the closeness of two or more measurements to each other.

Imprecision is the standard deviation (SD) or coefficient of variation (%CV) of analytic results in a set of replicate measurements. Fresh whole blood specimens used to verify imprecision specifications should have mean values that fall within the range tested in the following table and should not display any Suspect Parameter flags for the measurand (parameter) studied.

Whole blood sample repeatability requirements (calibration parameters)

Parameters	Measuring range Whole Blood	Repeatability CV%
WBC	$\geq 4.0 \times 10^9/L$	$\leq 2.5\%$
RBC	$\geq 3.5 \times 10^{12}/L$	$\leq 1.5\%$
HGB	(110-180)g/L	$\leq 1.5\%$
MCV	(70-120)fL	$\leq 1.0\%$
PLT	$\geq 150 \times 10^9/L$	$\leq 4.0\%$

Carryover

“Carryover” in laboratory testing, is defined as “the contamination of a specimen by the previous one”. Carryover testing is performed to help to prove or disprove carryover from the sample probe in clinical laboratory testing.

Carryover requirements for blood sample analysis

Measurement Parameters	Carryover
WBC	≤ 0.5
RBC	≤ 0.5
HGB	$\leq 0.6\%$
HCT	$\leq 0.5\%$
PLT	$\leq 1.0\%$

***Appendix

Performance Verification results

PERFORMANCE QUALIFICATION

SYSTEM CERTIFICATION

Study data has determined that the system described in this document meets all criteria outlined in this Performance Qualification protocol. All exceptional conditions if any have been addressed. The system is ready for specified usage.

Protocol performed by: Mindray Representative


Name: ARUNKUMAR PJ

Title: ASST APPLICATION MANAGER

Customer Authorization:

Name:

Title:



Representative Signature

Date: 14.09.2022

Customer Signature

Date:

CALIBRATION DATA BC 5130 TR-21006640

Lot No.	PLUS0722	Exp. Date	05/08/2022	Exp. Date	05/08/2022	Exp. Date	05/08/2022
Operator	Admin	Completion time of calibration	27/07/2022	Completion time of calibration	27/07/2022	Completion time of calibration	27/07/2022
	Select	RBC		HGB		MCV	PLT
		10 ⁶ /uL		g/dL		fl	10 ³ /uL
Target							
1	Yes	9.04	4.57	13.5	89.4	254	
2	No	8.94	4.43	13.3	90.4	254	
3	Yes	8.75	4.39	13.3	90.3	244	
4	Yes	8.79	4.33	13.3	90.5	242	
5	Yes	8.89	4.4	13.3	90.1	246	
6	Yes	8.96	4.42	13.4	90.0	246	
7	Yes	8.78	4.38	13.3	90.0	244	
8	No	8.8	4.43	13.4	90.1	237	
9	Yes	8.94	4.4	13.3	90.2	247	
10	Yes	8.85	4.4	13.3	90.6	235	
Mean		8.73	4.34	13.1	90.6	247	
CV (%)		8.84	4.39	13.3	90.3	244	
New Factor (%)	1	0.8		0.6	0.3	2.2	
Old Factor (%)	102.23	104.08		101.45	99.03	104.04	
	100.00	100.00		100.00	100.00	100.00	