



Cell : 9550411704

LAKSHMI LAB INSTRUMENTS

308, Rajarajeshwari Nilayam, St.No.1, Sri Rama Hills Colony,
Mansoorabad, L.B.Nagar, Hyderabad - 500068. Email : lakshmilabinstruments@gmail.com

CALIBRATION REPORT

CERTIFICATE NUMBER	2022-03/01-04
DATE OF CALIBRATION	01/03/2022
DUE DATE FOR CALIBRATION	28/02/2023

CLIENT DETAILS

NAME	AMERI DIAGNOSTICS & SCAN CENTER
ADDRESS	# 8-4-68/7/19, BESIDE AMBER GARDEN FUNCTION HALL, BANDLAGUDA, MAILARDEVAPALLY, RAJENDRA NAGAR, R.R DIST, TELANGANA 500005

INSTRUMENT DETAILS

NAME OF INSTRUMENT	LABORATORY CENTRIFUGE
MANUFACTURER	REMI
RANGE	UP TO 6000RPM
MODEL	C854/8
SERIAL / ID NUMBER	ZEGN-22908

DETAILS OF STANDARD INSTRUMENT USED

NAME	TACHOMETER
MASTER ID NUMBER	SL2108PL0200-01

TEST RESULTS

SLNO	DISPLAYED READING	MEASURED READING	ERROR
1	1800RPM	1820RPM	20RPM
2	2800RPM	2830RPM	30RPM
3	4500RPM	4580RPM	80RPM

