



Cell : 9550411704

# LAKSHMI LAB INSTRUMENTS

# 308, Rajarajeshwari Nilayam, St.No.1, Sri Rama Hills Colony,  
Mansoorabad, L.B.Nagar, Hyderabad - 500068. Email : lakshmilabinstruments@gmail.com

## CALIBRATION REPORT

CERTIFICATE NUMBER	2022-03/01-05
DATE OF CALIBRATION	01/03/2022
DUE DATE FOR CALIBRATION	28/02/2023

### CLIENT DETAILS

NAME	AMERI DIAGNOSTICS & SCAN CENTER
ADDRESS	# 8-4-68/7/19, BESIDE AMBER GARDEN FUNCTION HALL, BANDLAGUDA, MAILARDEVAPALLY, RAJENDRA NAGAR, R.R DIST, TELANGANA 500005

### INSTRUMENT DETAILS

NAME OF INSTRUMENT	THERMOMETER
MANUFACTURER	EUROLAB
RANGE	-50 TO 300 °C
MODEL	ST9283B
SERIAL / ID NUMBER	NA

### DETAILS OF STANDARD INSTRUMENT USED

NAME	UNIVERSAL TEMPERATURE INDICATOR
MASTER ID NUMBER	SL2108TL0441-001

### TEST RESULTS

SLNO	DISPLAYED READING	MEASURED READING	ERROR
1	50.0°C	50.5°C	0.1°C
2	100.0°C	101.0°C	1.0°C
3	150.0°C	151.0°C	1.0°C
4	250.0°C	252.0°C	2.0°C
5	-20.0°C	-20.2°C	0.2°C
6	-40.0°C	-39.0°C	1.0°C

