

CALIBRATION CERTIFICATE

EQN/FT/7.8/01

Page 1 of 1

CRF No.	: EQN/CRF/2209034	Date of Calibration	: 19-Sep-22
Certificate No	: 2022-23/EQN/2209034-13	Recom. Due Date	: 18-Sep-23
ULR No.	: CC276022000012732F		

Customer Details	Calibrated at	: Onsite
M/s. SRI KRISHNA CLINICAL LABORATORY	Date of Receipt	: 19-Sep-22
No : 2 Karungalpatti, 2nd street, Gugai,	Cond. On Receipt	: Satisfactory
Salem- 636 006.	Date of Issue	: 22-Sep-22

DETAILS OF UUC :			
Description	: Dry Bath Incubator	Identification No.	: SCL-004
Range	: Amb to 70°C	Serial No.	: 4948
Least Count	: 0.1°C	Make	: Labtech
Cal./Working Range	: 37°C	Model / Type	: DB-11
Location	: Lab	Accuracy	: As Per Manual

DETAILS OF STANDARD USED :			
Master Name	Lab ID No. / Sr No.	Valid upto	Traceability
Portable Calibrator	INS/ET/02/20786	04-Dec-22	EQN, Chennai
RTD	INS/TM/29	02-Aug-23	EQN, Chennai




CP No.	: EQN/CP/TL-03
Environmental Details	: Temperature: 25±5°C Relative Humidity: 45-70 % RH

THERMAL CALIBRATION - (Temperature)

CALIBRATION RESULT				
Sl.No.	TEST READING	STD READING	ERROR	UNCERTAINTY
	(°C)	(°C)	(°C)	± (°C)
1	37.6	37.23	0.37	0.40

Remark :

1. Reading given was the average of 3 Measurements.
2. The reported Expanded Uncertainty is calculated at 95 % C.L Approx with coverage factor k=2
3. Temperature scale:ITS 90

Issued by :  Engineer - Calibration		Authorized by :  K.YOGANATHAN
--	---	--

EQUATION CALIBRATION SERVICES PRIVATE LIMITED