



Cell : 9550411704

LAKSHMI LAB INSTRUMENTS

308, Rajarajeshwari Nilayam, St.No.1, Sri Rama Hills Colony,
Mansoorabad, L.B.Nagar, Hyderabad - 500068. Email : lakshmilabinstruments@gmail.com

CALIBRATION REPORT

CERTIFICATE NUMBER	2022-03/01-02
DATE OF CALIBRATION	01/03/2022
DUE DATE FOR CALIBRATION	28/02/2023

CLIENT DETAILS

NAME	AMERI DIAGNOSTICS & SCAN CENTER
ADDRESS	# 8-4-68/7/19, BESIDE AMBER GARDEN FUNCTION HALL, BANDLAGUDA, MAILARDEVAPALLY, RAJENDRA NAGAR, R.R DIST, TELANGANA 500005

INSTRUMENT DETAILS

NAME OF INSTRUMENT	MICRO PIPPETTE
MANUFACTURER	DRAGON LAB
RANGE	10-100 μ L
MODEL	NA
SERIAL / ID NUMBER	YE3K143876

TEST RESULTS

SET VALUE	RECORDED VALUE
10 μ L	0.015g
25 μ L	0.026g
50 μ L	0.059g
75 μ L	0.076g
100 μ L	0.152g

