



Cell : 9550411704

LAKSHMI LAB INSTRUMENTS

308, Rajarajeshwari Nilayam, St.No.1, Sri Rama Hills Colony,
Mansoorabad, L.B.Nagar, Hyderabad - 500068. Email : lakshmilabinstruments@gmail.com

CALIBRATION REPORT

CERTIFICATE NUMBER	2022-03/01-03
DATE OF CALIBRATION	01/03/2022
DUE DATE FOR CALIBRATION	28/02/2023

CLIENT DETAILS

NAME	AMERI DIAGNOSTICS & SCAN CENTER
ADDRESS	# 8-4-68/7/19, BESIDE AMBER GARDEN FUNCTION HALL, BANDLAGUDA, MAILARDEVAPALLY, RAJENDRA NAGAR, R.R DIST, TELANGANA 500005

INSTRUMENT DETAILS

NAME OF INSTRUMENT	MICRO PIPETTE
MANUFACTURER	DRAGON LAB
RANGE	5-50 μ L
MODEL	NA
SERIAL / ID NUMBER	YE312AS0147570

TEST RESULTS

SET VALUE	RECORDED VALUE
5 μ L	0.005g
10 μ L	0.010g
25 μ L	0.024g
50 μ L	0.048g

