

**RML EQAS CORRECTIVE &
PREVENTIVE ACTION**

Name of the department: Hematology

EQAS Programme : RML EQAS

Cycle / Sample No : 11

Month and year : AUGUST - 2022

Nature of Non conformance:

Hemoglobin and RBC SDI : $\geq \pm 2.0$

Corrective and preventive action taken:

Hemoglobin and RBC Reagent stability, Expiry status and internal quality Control values were verified on the day (12/07/22) we processed our EQAS sample and the QC value is within the normal range only. It may be due to Random Error.

Review in the next EQAS sample.

Date:

21/8/22



Approved by:

**Dr Mohd Khaleel Msc.Ph.D
Head-Clinical Laboratory**