

CERTIFICATE OF CALIBRATION

Customer Name	Govt: Taluk Head Quarters Hospital	Certificate No.	THP111
Customer Address	Panamkuttymala, Punalur, Kerala 691305	Date of Calibration	22/03/2022
		Valid Upto	21/03/2023
		Date of Issue	23/03/2022

DUC* Details			
DUC* Name	CENTRIFUGE	S No.	8PLC-2999
Make/Model No.	REMI	Location	ICTC LAB
ID No.	..	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	DIGITAL TACHO METER	THERMO HYGROMETER
Make	METRAVI	HTC-2
Model/SR No.	NCTM 1000/190518478	...
Equipment Id	CHPL/MT-KL/TAM/002	
Calibration certificate No	UCSL/07-21/201_06	UCSL/07-21/201-01
Calibration Validity	20/07/2022	20/07/2022
Calibration By	UNIQUE CALIBRATIONS	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	27°C	Relative Humidity	54%

Calibration Results				
SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE	STATUS
1	500	512	±5%	PASS
2	1000	1028	±5%	PASS
3	2500	2548	±5%	PASS

CALIBRATION STATUS	PASS	✓	FAIL	
--------------------	------	---	------	--

REMARKS:- *DUC- DEVICE UNDER CALIBRATION

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

JAYADEV K S

Issued and Approved By

SARATH K B
TECHNICAL MANAGER

