

CERTIFICATE OF CALIBRATION

Customer Name	Govt:Taluk Head Quarters Hospital	Certificate No.	THP113
CUSTOMER Address	Panamkuttymala, Punalur, Kerala 691305	Date of Calibration	22/03/2022
		Valid Upto	21/03/2023
		Date of Issue	23/03/2022

DUC* Details			
DUC* Name	REFRIGERATOR	S No.	..
Make/Model No.	ELECTROLUX	Location	ICTC LAB
ID No.	..	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	THERMO HYGROMETER	ELECTRICAL SAFETY ANALYZER	INFRARED THERMOMETER
Make	HTC-2	FLUKE	FLUKE
Model/SR No.	---	ESA 612/2959012	59 MAX/49958584WS
Equipment Id		CHPL/MT-KL/ESA/002	CHPL/MT-KL/IR/001
Calibration certificate No	UCSL/07-21/201-01	ET/220124A/01	UCSL/07-21/201_07
Calibration Validity	20/07/2022	25/01/2023	20/07/2022
Calibration By	UNIQUE CALIBRATION SOLUTIONS	STIC	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET VALUE(°C)	MEASURED VALUE (°C)	TOLERANCE	STATUS
1	4	4	±5%	PASS
2	4	4	±5%	PASS
3	4	4	±5%	PASS

CALIBRATION STATUS	PASS	✓	FAIL
--------------------	------	---	------

REMARKS:- *DUC – DEVICE UNDER CALIBRATION

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
JAYADEV K S

Issued and Approved By
SARATH K B
TECHNICAL
MANAGER



Skb