



Ref : TBMNZ-1/2223/NKC866
Dated : 29-07-2022

Calibration Certificate

Customer Name : Pathkind Labs ,Tilak Nagar, Delhi.

Model : Easy Lyte Plus

Serial No. : 61866CNKC

Calibration Date : 29th July 2022

Calibration Due Date : 28th July 2023

This is to certify that we have calibrated the Fully Automated Electrolyte Analyser , Model : Easy Lyte Plus bearing Sr. No. 61866CNKC and on inspection of the instrument it is observed that the results are well within the range and instrument is working fine .

Thanking you,

For Transasia Bio-Medicals Ltd.

Sunil Gautam
Service Engineer

TECHNICAL SERVICE REPORT No. **1441885**

DATE: **29/07/2022**

| CUSTOMER DETAILS | | INSTRUMENT DETAILS | | SERVICE STATUS | |
|---|-----------------------|--------------------------|--|---|------------------|
| NAME: Dain Kund 200 | | MODEL: MAKIC1 | | <input type="checkbox"/> WARRANTY <input type="checkbox"/> R&R <input checked="" type="checkbox"/> AMC <input type="checkbox"/> CMC <input type="checkbox"/> CHARGED CALL | |
| ADDRESS: Tilak Nagar Delhi | | SR. NO.: 61866CAK | | TYPE OF CALL <input type="checkbox"/> INSTALLATION <input checked="" type="checkbox"/> P.M. VISIT <input type="checkbox"/> I <input type="checkbox"/> II <input type="checkbox"/> III <input type="checkbox"/> IV <input type="checkbox"/> APPLICATION SUPPORT <input type="checkbox"/> BREAKDOWN | |
| | | CALL DETAILS | | | |
| | | COMPLAINT RECD. | DATE | TIME | |
| | | RESPONSE | 29/7/2022 | | |
| TEL NO.: | | JOB COMPLETED | 29/7/2022 | | |
| NAME OF THE OPERATOR: Mr. Rajesh | | RESPONSE TIME | | | DOWN TIME: |
| PROBLEM REPORTED: Preventive maintenance | | TRAVEL TIME | | | COUNTER READING: |
| OBSERVATIONS: Pr. Bulb (membrane and tubing required for scheduled change) | | | | | |
| ACTION TAKEN: Clean the analyzer. Carried the maintenance. checked electrodes working ok | | | | | |
| SITE CONDITION: LINE-NEUTRAL VOLT.: 229V NEUTRAL-EARTH VOLT.: 0V LINE-EARTH VOLT.: 231V | | | | | |
| BRAND OF REAGENT USED: | | | TO BE FILLED IN BY CUSTOMER | | |
| <input type="checkbox"/> FOLLOWING PARTS HAVE BEEN REPLACED <input checked="" type="checkbox"/> FOLLOWING PARTS NEED TO BE REPLACED. PLEASE APPROVE | | | <input type="checkbox"/> PREVENTIVE MAINTENANCE CARRIED OUT SATISFACTORILY. <input type="checkbox"/> FAULT RECTIFIED & INSTRUMENT IS WORKING SATISFACTORILY. <input type="checkbox"/> WE HEREBY APPROVE RS. _____ FOR PARTS <input type="checkbox"/> COMMENTS (IF ANY): | | |
| NO. | DESCRIPTION | QTY. | COST | TOTAL | |
| 160283 | Membrane assly | 01 | 3000.00 | 3000/- | |
| 181655 | Tubing kit | 01 | 2700.00 | 2700/- | |
| | | | | 5700.00 + 18% | |
| TOTAL Rs. | | | | Rs. 6726.00 | |
| INVOICE NO.: | | DATE: | | BRANCH H. O. | |
| FOLLOW-UP ACTION (Required if any): | | | | | |
| ENGINEER'S/APPLICATION SPECIALIST'S SIGNATURE: Smit | | | | RECEIVED ON: | |
| TIME: NAME: Smit | | | | CHECKED BY: | |
| | | | | JOB CARD NO.: | |
| <small>NOTE: Parts replaced are chargeable except during warranty period. Consumables like printer head, lamp, tubing, paper rolls etc. & breakable parts are not covered by warranty and hence are chargeable. Parts replaced due to negligence in operation will also be charged in every case.</small> | | | | | |
| AT TRANSASIA, CUSTOMER SATISFACTION IS OUR PRIME CONCERN. IN CASE YOU HAVE ANY SUGGESTIONS PLEASE CONTACT: GENERAL MANAGER (TECHNICAL SERVICE), MUMBAI. TEL.: 4030 9000 | | | | | |

TRANSASIA BIO-MEDICALS LTD.

Head Office : Transasia House, 8, Chandivali Studio Road, Andheri (E), Mumbai - 400 072. Tel. : (022) 4030 9000 Fax : (022) 2857 3030
 Mumbai : Toll Free No. 1800 103 8226, Whatsapp No. : 8451048434, SMS No. : 9212433444
 Indore : Tel. : (0731) 403 3948 Chennai : Tel. : (044) 2822 7149 Kolkata : Tel. : (033) 2215 7839 Fax : 2215 8035
 Lucknow : Tel. : (0522) 443 0900 Pune : Tel. : (020) 2616 2658 Dehi : Tel. : (011) 2578 5451/2573 2223
 Chandigarh : Tel. : (0172) 464 3837 Hyderabad : Tel. : (040) 4430 9000 Ahmedabad : Tel./Fax : (079) 02640 7030/4032 1903
 Jaipur : Tel. : 93092 63147 Guwahati : Tel. : 91369 50480 Bangalore : Tel./Fax : (080) 2556 8044
 Patna : Tel. : 77669 00553 Chuttack : Tel. : (0671) 232 4555
 Kochi : Tel. : (0484) 402 0511 Hubli : Tel. : (0836) 485 0900

DOC NO : SC00-406 / ISS - 6

www.transasia.co.in

