



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL LABORATORY CENTRIFUGE

Certificate No:	22900004214	Issued On:	16/09/2022
ULR No:	CC214622900004214F	Calibration Date:	14/09/2022
Job Identification No:	ICS/C/AS/09/4214	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:

M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	RPM Measurement for Digital Laboratory Centrifuge		
Make/Model	REMI / R-8C Plus	Visual Inspection	Ok
Serial No.	---	I.D. No.	CEN-01
Range	Up to 6000 RPM	Least Count	10 RPM
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

USED EQUIPMENT DETAILS

Not Applicable

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-RPM01
(Comparison Method)

RESULTS: Mechanical Calibration MEASUREMENT: (in RPM)

CENTRIFUGE' READING (AVERAGE)	CORRECTED VALUE OF TACHOMETER READING (AVG.)	UNCERTAINTY AT APPROX 95% C.L AND COVERAGE FACTOR K=2
1220	1221.3	± 6
2130	2132.8	± 6
2720	2725.4	± 7
3510	3518.9	± 7
4510	4522.3	± 8

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Authorized Signatory
Operation Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL LABORATORY CENTRIFUGE

Certificate No:	22900004213	Issued On:	16/09/2022
ULR No:	CC214622900004213F	Calibration Date:	14/09/2022
Job Identification No:	ICS/C/AS/09/4213	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:

M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	RPM Measurement for Digital Laboratory Centrifuge		
Make/Model	REMI / R-8C Plus	Visual Inspection	Ok
Serial No.	ZHFN-18294	I.D. No.	CEN-02
Range	Up to 6000 RPM	Least Count	10 RPM
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

USED EQUIPMENT DETAILS

Not Applicable

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-RPM01
(Comparison Method)

RESULTS: Mechanical Calibration MEASUREMENT: (in RPM)


CENTRIFUGE' READING (AVERAGE)	CORRECTED VALUE OF TACHOMETER READING (AVG.)	UNCERTAINTY AT APPROX 95% C.I. AND COVERAGE FACTOR K=?
1220	1220.8	± 6
2130	2132.2	± 6
2720	2724.6	± 7
3510	3517.3	± 7
4510	4521.6	± 8

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by  RAJESH KAPOOR
Operation-Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL V.D.R.L. SHAKER

Certificate No:	22900004212	Issued On:	16/09/2022
ULR No:	CC214622900004212F	Calibration Date:	14/09/2022
Job Identification No:	ICS/C/AS/09/4212	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:	M/s Pathkind Diagnostics Pvt. Ltd., Deen Dayal Hospital, Complex Plot No-310, C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004
------------------------	--

EQUIPMENT DESCRIPTION

Name	RPM Measurement for Digital V.D.R.L. Shaker		
Make/Model	REMI / RS-12 R	Visual Inspection	Ok
Serial No.	ZHCS-08774	I.D. No.	---
Range	Upto 180 RPM	Least Count	1 RPM
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

USED EQUIPMENT DETAILS: Not Applicable

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Comparison Method)

RESULTS: Mechanical Calibration
MEASUREMENT: (in RPM)


CENTRIFUGE' READING (AVERAGE)	CORRECTED VALUE OF TACHOMETER READING (AVG.)	UNCERTAINTY AT APPROX 95% C.L AND COVERAGE FACTOR K=2
40	40.3	± 4
80	80.6	± 4
120	120.8	± 4
150	151.1	± 5
170	171.4	± 5

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by  **RAJESH KAPOOR**
Authorized Signatory
Operation-Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechcs.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF CYCLO MIXER

Certificate No:	22900004211	Issued On:	16/09/2022
ULR No:	CC214622900004211F	Calibration Date:	14/09/2022
Job Identification No:	ICS/C/AS/09/4211	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:	M/s Pathkind Diagnostics Pvt. Ltd., Deen Dayal Hospital, Complex Plot No-310, C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004
------------------------	--

EQUIPMENT DESCRIPTION

Name	RPM Measurement for Cyclo Mixer		
Make/Model	Remi / CM-101	Visual Inspection	Ok
Serial No.	ZHCS-10480	I.D. No.	---
Range	Upto 2800 RPM	Least Count	N/A
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

USED EQUIPMENT DETAILS	Not Applicable
-------------------------------	----------------

PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Measurement Method)
--	--

RESULTS: Mechanical Calibration

MEASUREMENT: (in RPM)

U.U.C. SPEED KNOB' POSITION	CORRECTED VALUE OF TACHOMETER READING(Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2
At Minimum	902.1	± 3
At Middle	1656.6	± 3
At Maximum	2699.8	± 3

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Authorised Signatory
Operation Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
Phone / Fax : (011) 22453259, 22434362, 9599482307
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF GLASS THERMOMETER

Certificate No:	22100014630	Issued On:	16/09/2022
ULR No:	CC214622100014630F	Calibration Date:	14/09/2022
Job Identification No:	ICS/C/T/09/14630	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:

M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	Glass Thermometer		
Make/Model	Dyna	Visual Inspection	Ok
Serial No.	---	I.D. No.	GT-01
Range	-50 to 50°C	Least Count	0.5°C
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 5.0° C	Humidity	50 ± 15 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000002	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

USED EQUIPMENT DETAILS

Methanol Bath & Water Bath

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS-/CAL/SOP-TH01, ITS-90
(Comparison Method)

RESULTS: Thermal Calibration (On Working Range)

U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
-20.0	-20.05	± 0.35
00.0	-00.01	± 0.35
10.0	09.98	± 0.35
25.0	24.96	± 0.35
40.0	39.92	± 0.41

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

*** END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Authorised Signatory
Calibration Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF GLASS THERMOMETER

Certificate No:	22100014629	Issued On:	16/09/2022
ULR No:	CC214622100014629F	Calibration Date:	14/09/2022
Job Identification No:	ICS/C/T/09/14629	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:

M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	Glass Thermometer		
Make/Model	Dyna	Visual Inspection	Ok
Serial No.	---	I.D. No.	GT-02
Range	-50 to 50°C	Least Count	0.5°C
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 5.0° C	Humidity	50 ± 15 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000002	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

USED EQUIPMENT DETAILS

Methanol Bath & Water Bath

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS-/CAL/SOP-TH01, ITS-90
(Comparison Method)

RESULTS: Thermal Calibration (On Working Range)

U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
-20.0	-20.06	± 0.35
00.0	-00.02	± 0.35
10.0	09.97	± 0.35
25.0	24.95	± 0.35
40.0	39.91	± 0.41

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

*** END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Authorised Signatory
Operation Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF GLASS THERMOMETER

Certificate No:	22100014628	Issued On:	16/09/2022
ULR No:	CC214622100014628F	Calibration Date:	14/09/2022
Job Identification No:	ICS/C/T/09/14628	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:	M/s Pathkind Diagnostics Pvt. Ltd., Deen Dayal Hospital, Complex Plot No-310, C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004
------------------------	---

EQUIPMENT DESCRIPTION

Name	Glass Thermometer		
Make/Model	Dyna	Visual Inspection	OK
Serial No.	---	I.D. No.	GT-03
Range	-50 to 50°C	Least Count	0.5°C
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 5.0 °C	Humidity	50 ± 15 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000002	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

USED EQUIPMENT DETAILS Methanol Bath & Water Bath

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS-/CAL/SOP-TH01, ITS-90 (Comparison Method)

RESULTS: Thermal Calibration
(On Working Range)

U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
-20.0	-20.06	± 0.35
00.0	-00.02	± 0.35
10.0	09.97	± 0.35
25.0	24.95	± 0.35
40.0	39.91	± 0.41

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

*** END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Operation-Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF GLASS THERMOMETER

Certificate No:	22100014654	Issued On:	16/09/2022
ULR No:	CC214622100014654F	Calibration Date:	14/09/2022
Job Identification No:	ICS/C/T/09/14654	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:

M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	Glass Thermometer		
Make/Model	Dyna	Visual Inspection	Ok
Serial No.	---	I.D. No.	GT-04
Range	-50 to 50°C	Least Count	0.5°C
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 5.0° C	Humidity	50 ± 15 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000002	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

USED EQUIPMENT DETAILS

Methanol Bath & Water Bath

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS-/CAL/SOP-TH01, ITS-90
(Comparison Method)

RESULTS: Thermal Calibration (On Working Range)

U.U.C. READING (in °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
-20.0	-19.91	± 0.35
00.0	00.03	± 0.35
10.0	10.06	± 0.35
25.0	25.11	± 0.35
40.0	40.17	± 0.41

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

*** END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAROOR
Authorised Signatory
Operation-Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
Phone / Fax : (011) 22453259, 22434362, 9599482307
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL ELECTRONIC TIMER

Certificate No:	22500004530	Issued On	16/09/2022
ULR No:	CC214622500004530F	Calibration Date	14/09/2022
Job Identification No:	ICS/C/E-T/09/4530	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:	M/s Pathkind Diagnostics Pvt. Ltd., Deen Dayal Hospital, Complex Plot No-310, C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004
------------------------	--

EQUIPMENT DESCRIPTION

Name	Digital Timer		
Make/Model	Gem Loudstar	Visual Inspection	Ok
Serial No.	---	I.D. No.	DT-01
Range	Up to 99 hr. 59 min. 59 sec.	Least Count	1 sec
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make/Model	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Stop Watch	Taksun	22500000003	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

USED EQUIPMENT DETAILS

N/A

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-ET12
(Comparison Method)

RESULTS: Electro Technical Calibration

U.U.C. TIME SET	STD. OBSERVED READING (AVERAGE)	Uncertainty at approx 95% C.L and coverage factor k=2
01 Min. 00 s	01 Min. 00.12 s	± 1.2 s
05 Min. 00 s	05 Min. 05.24 s	± 1.2 s
15 Min. 00 s	15 Min. 00.39 s	± 1.2 s
30 Min. 00 s	30 Min. 01 s	± 1.2 s


REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

*** END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by 
RAJESH KAPOOR
Authorised Signatory
Operation-Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL ELECTRONIC TIMER

Certificate No:	22500004531	Issued On	16/09/2022
ULR No:	CC214622500004531F	Calibration Date	14/09/2022
Job Identification No:	ICS/C/E-T/09/4531	Next Calibration Date:	13/09/2023
Ref. No:	SRF, Dated -12/09/2022		

CALIBRATED FOR:

M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	Digital Timer		
Make/Model	---	Visual Inspection	Ok
Serial No.	---	I.D. No.	DT-02
Range	Up to 23 hr. 59 min. 59 sec.	Least Count	1 sec
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make/Model	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Stop Watch	Taksun	22500000003	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

USED EQUIPMENT DETAILS

N/A

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-ET12
(Comparison Method)

RESULTS: Electro Technical Calibration

U.U.C. TIME SET	STD. OBSERVED READING (AVERAGE)	Uncertainty at approx 95% C.L and coverage factor k=2
01 Min. 00 s	01 Min. 00.17 s	± 1.2 s
05 Min. 00 s	05 Min. 05.32 s	± 1.2 s
15 Min. 00 s	15 Min. 00.48 s	± 1.2 s
30 Min. 00 s	30 Min. 01 s	± 1.2 s

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

*** END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Authorised Signatory
Operation-Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
Phone / Fax : (011) 22453259, 22434362, 9599482307
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL THERMO-HYGROMETER

Calibration Certificate No.	22100014655	Issued On	16/09/2022
ULR No.	CC214622100014655F	Calibration Date	14/09/2022
Job Identification No.	ICS/C/T/09/14655	Next Due Date	13/09/2023
Ref. No.	SRF, Dated - 12/09/2022		

CALIBRATED FOR:

M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	Digital Thermo-Hygrometer		
Make/Model	HTC-1	Visual Inspection	Ok
Serial No.	---	I.D. No.	DTH-01
Range	Up to 60 °C / Up to 99 % (RH)	Least Count	0.1°C / 1 % RH
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator	Yudian	ETL/2113/22/C/F/061	FCRI, Palakkad	02-03-2022 to 04-03-2022	04-03-2023
Temp. & Hum. Ind. with Sensor	Rotronic	ETL/1703/21/C/F/506	FCRI, Kerala	30/12/2021 - 04-01-2022	04/01/2023

USED EQUIPMENT DETAILS: Environmental Chamber

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-TH06 (ITS-90) (Comparison Method)

RESULTS: Thermal Calibration
(On Working Range)

U.U.C. Reading		Corrected Value of Standard		Uncertainty at approx 95% C.L and coverage factor k=2	
Temperature	Humidity	Temperature	Humidity	Temperature	Humidity
20.0 °C	40 % RH	20.038°C	39.78 %RH	± 0.32°C	± 1.6 % RH
25.0 °C	60 % RH	25.085°C	59.18 %RH	± 0.32°C	± 1.6 % RH
30.0 °C	75 % RH	30.135°C	73.49 %RH	± 0.32°C	± 1.6 % RH
40.0 °C	90 % RH	40.176°C	88.12 %RH	± 0.32°C	± 1.6 % RH


REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by


RAKESH KAPOOR
Operation-Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
Phone / Fax : (011) 22453259, 22434362, 9599482307
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL THERMO-HYGROMETER

Calibration Certificate No.	22100014656	Issued On	16/09/2022
ULR No.	CC214622100014656F	Calibration Date	14/09/2022
Job Identification No.	ICS/C/T/09/14656	Next Due Date	13/09/2023
Ref. No.	SRF, Dated - 12/09/2022		

CALIBRATED FOR:

M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	Digital Thermo-Hygrometer		
Make/Model	HTC-1	Visual Inspection	Ok
Serial No.	---	I.D. No.	DTH-02
Range	Up to 60 °C / Up to 99 % (RH)	Least Count	0.1°C / 1 % RH
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator	Yudian	ETL/2113/22/C/F/061	FCRI, Palakkad	02-03-2022 to 04-03-2022	04-03-2023
Temp. & Hum. Ind. with Sensor	Rotronic	ETL/1703/21/C/F/506	FCRI, Kerala	30/12/2021 - 04-01-2022	04/01/2023

USED EQUIPMENT DETAILS

Environmental Chamber

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-TH06 (ITS-90)
(Comparison Method)

RESULTS: Thermal Calibration (On Working Range)

U.U.C. Reading		Corrected Value of Standard		Uncertainty at approx 95% C.L and coverage factor k=2	
Temperature	Humidity	Temperature	Humidity	Temperature	Humidity
20.0 °C	40 % RH	19.972°C	39.56 %RH	± 0.32°C	± 1.6 % RH
25.0 °C	60 % RH	24.935°C	59.05 %RH	± 0.32°C	± 1.6 % RH
30.0 °C	75 % RH	29.898°C	73.55 %RH	± 0.32°C	± 1.6 % RH
40.0 °C	90 % RH	39.861°C	88.05 %RH	± 0.32°C	± 1.6 % RH

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by 
RAJESH KAPOOR
Authorised Signatory
Operations Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL THERMO-HYGROMETER

Calibration Certificate No.	22100014657	Issued On	16/09/2022
ULR No.	CC214622100014657F	Calibration Date	14/09/2022
Job Identification No.	ICS/C/T/09/14657	Next Due Date	13/09/2023
Ref. No.	SRF, Dated - 12/09/2022		

CALIBRATED FOR:	M/s Pathkind Diagnostics Pvt. Ltd., Deen Dayal Hospital, Complex Plot No-310, C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004
------------------------	---

EQUIPMENT DESCRIPTION

Name	Digital Thermo-Hygrometer		
Make/Model	HTC-1	Visual Inspection	Ok
Serial No.	---	I.D. No.	DTH-03
Range	Up to 60 °C / Up to 99 % (RH)	Least Count	0.1°C / 1 % RH
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator	Yudian	ETL/2113/22/C/F/061	FCRI, Palakkad	02-03-2022 to 04-03-2022	04-03-2023
Temp. & Hum. Ind. with Sensor	Rotronic	ETL/1703/21/C/F/506	FCRI, Kerala	30/12/2021 - 04-01-2022	04/01/2023

USED EQUIPMENT DETAILS	Environmental Chamber
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-TH06 (ITS-90) (Comparison Method)

RESULTS: Thermal Calibration (On Working Range)


U.U.C. Reading		Corrected Value of Standard		Uncertainty at approx 95% C.L and coverage factor k=2	
Temperature	Humidity	Temperature	Humidity	Temperature	Humidity
20.0 °C	40 % RH	19.964°C	39.53 %RH	± 0.32°C	± 1.6 % RH
25.0 °C	60 % RH	24.921°C	59.09 %RH	± 0.32°C	± 1.6 % RH
30.0 °C	75 % RH	29.881°C	73.47 %RH	± 0.32°C	± 1.6 % RH
40.0 °C	90 % RH	39.834°C	87.95 %RH	± 0.32°C	± 1.6 % RH

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 3. Total liability of our Organisation is limited to the invoiced amount.
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by 
RAJESH KAPOOR
 Operation-Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL THERMO-HYGROMETER

Calibration Certificate No.	22100014658	Issued On	16/09/2022
ULR No.	CC214622100014658F	Calibration Date	14/09/2022
Job Identification No.	ICS/C/T/09/14658	Next Due Date	13/09/2023
Ref. No.	SRF, Dated - 12/09/2022		

CALIBRATED FOR:

M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	Digital Thermo-Hygrometer		
Make/Model	HTC-1	Visual Inspection	Ok
Serial No.	---	I.D. No.	DTH-04
Range	Up to 60 °C / Up to 99 % (RH)	Least Count	0.1°C / 1 % RH
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator	Yudian	ETL/2113/22/C/F/061	FCRI, Palakkad	02-03-2022 to 04-03-2022	04-03-2023
Temp. & Hum. Ind. with Sensor	Rotronic	ETL/1703/21/C/F/506	FCRI, Kerala	30/12/2021 - 04-01-2022	04/01/2023

USED EQUIPMENT DETAILS: Environmental Chamber

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-TH06 (ITS-90) (Comparison Method)

RESULTS: Thermal Calibration
(On Working Range)


U.U.C. Reading		Corrected Value of Standard		Uncertainty at approx 95% C.L and coverage factor k=2	
Temperature	Humidity	Temperature	Humidity	Temperature	Humidity
20.0 °C	40 % RH	20.041°C	40.39 %RH	± 0.32°C	± 1.6 % RH
25.0 °C	60 % RH	25.078°C	60.82 %RH	± 0.32°C	± 1.6 % RH
30.0 °C	75 % RH	30.121°C	76.24 %RH	± 0.32°C	± 1.6 % RH
40.0 °C	90 % RH	40.174°C	91.74 %RH	± 0.32°C	± 1.6 % RH

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services


Checked by


Authorised Signatory
Operation-Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechcs.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL TEMP. CONT. CUM INDICATOR WITH SENSOR

Calibration Certificate No.	22100014659	Issued On	16/09/2022
ULR No.	CC214622100014659F	Calibration Date	14/09/2022
Job Identification No.	ICS/C/T/09/14659	Next Due Date	13/09/2023
Ref. No.	SRF, Dated - 12/09/2022		

CALIBRATED FOR: M/s Pathkind Diagnostics Pvt. Ltd.,
Deen Dayal Hospital, Complex Plot No-310,
C.T.S. No-926, Shivaji Nagar (Bhamburda) Pune MA-411004

EQUIPMENT DESCRIPTION

Name	Digital Temperature Controller Cum Indicator with sensor Fitted in Incubator		
Make/Model	I-Therm / AI-7481	Visual Inspection	Ok
Serial No.	---	I.D. No.	INC-01
Range	Up to 110°C	Least Count	0.1°C
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 5.0 °C	Humidity	50 ± 15 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	22100000002	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

USED EQUIPMENT DETAILS

Water Bath

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS-/CAL/SOP-TH02, ITS-90
(Comparison Method)

RESULTS: Thermal Calibration

U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
30.0	29.87	± 0.30
35.0	34.82	± 0.30
37.0	36.80	± 0.30
45.0	44.79	± 0.30
60.0	59.65	± 0.42

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

*** END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
- The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 - Total liability of our Organisation is limited to the invoiced amount.
 - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services


Checked by


RAJESH KAPOOR
Authorised Signatory
Operation-Head