

# MK BEST CALIBRATION SERVICES



NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,  
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

## CERTIFICATE OF CALIBRATION

FF/7.8/01				Page No	1 of 1
URL No	CC334022000015793F	Date of Calibration	05.09.2022	Date of Receipt	05.09.2022
Certificate No	MKBL/22/09/1075-004	Recom. Due Date	04.09.2023	Date of Issue	05.09.2022

CUSTOMER INFORMATION	DETAILS OF UNIT UNDER CALIBRATION	
M/S. SRI BAALAJI CLINICAL LAB , VILLUPURAM.	Description	THERMO HYGROMETER
	Make / Model	HTC / HTC - 1
	Range	-10 to 50 ° C / 10% 99%RH
	Resolution	± 0.1 ° C / ± 1%RH
	ID / No	SBV/EQP/GEN/05
	Serial No	----
	Location	---
	Calibrated at	LAB

STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)				
S.No	Description	Id.No/Sl. No	Certificate No	Validity
01	Temperature & Humidity Indicator with sensor	MK/CAL-43 / HTI042K152714	CRMTL/01/422101234-A5	04.07.2023

ENVIRONMENTAL CONDITIONS		REFERENCE STANDARD	
Temperature	25 ± 4°C	Procedure No	MKBCS - TH - 09
Humidity	30 - 75 % RH	Condition of DUC	Good

### CALIBRATION RESULTS

1.THERMAL CALIBRATION TEMPERATURE				
Parameter / Range	Standard Reading ( ° C )	DUC Reading ( ° C )	Deviation ( ° C )	Expanded Uncertainty (±) (° C)
Temperature @ 50% RH	20.22	20.4	0.18	0.31
	30.12	30.5	0.38	
	40.08	40.5	0.42	
	50.14	50.6	0.46	

  

Parameter / Range	Standard Reading ( %RH )	UUC Reading ( %RH )	Deviation ( %RH )	Expanded Uncertainty (±) %RH
Humidity @ 25°C	30.3	31	-0.7	1.42
	50.3	52	-1.7	
	70.4	73	-2.6	
	90.4	93	-2.6	

- Remarks :
1. The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.
  2. The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
  3. The Recalibration Interval Should be Determined on the User Requirement.
  4. The Results Stated In This Certificate Relate Only to the Item Calibrated.
  5. The User Should Determine The Suitability Of The Instrument For Is Intended Use.
- x-x-x-x End Of Certificate x-x-x-x

Calibrated by  
  
 S.Chandra Bose  
 (Calibration Engineer)

Authorised By  
  
 L.Magesh  
 (MD/QM)