



## Rapid Diagnostic Pvt. Ltd.


A5, D. N. Shree Nivas Apartments (First Floor)

16, Venu Reddy Street,

Gundy, Chennai - 600 032

Phone 044 - 43534413, 98426 34354, 98412 47081

Email chennaioffice@rdgc.in



## Calibration Certificate

Customer Name	Sri Baalaji Clinical Lab
Address	No.87H, S.G.Puram, Main Road, (Opp.Santhi Theatre), Villupuram - 605602
Contact Person	Mr.B.Mohankumar
Contact No.	94432 90583, 99523 90583
Instrument	BioChemistry Analyzer
Model	STAR 20
Serial No.	ES1002010PN9434
Instrument Status	WORKING
Calibration Done On	14-07.2022
Next Calibration Due On	13-07-2023

### For Rapid Diagnostic Pvt. Ltd

Deputy Manager - Service



A.Rajeswaran

16-09-2022 15:51:09  
11.9397223  
79.4976515

QC Name : glucose  
Lot No. : 01  
Test Name: Glucose  
Abs. : 0.4400  
Result: 121.37  
Unit: mg/dl  
Mean: 114.00  
SD: 124.00  
Test/Date: 2022/09/16 15:48:39

QC Name : glucose  
Lot No. : 01  
Test Name: TGL  
Abs. : 0.2844  
Result: 159.48  
Unit: mg/dl  
Mean: 151.00  
SD: 164.00  
Test/Date: 2022/09/16 16:09:50

QC Name : 10C  
Lot No. : 01  
Test Name: CHOL  
Abs. : 0.2535  
Result: 120.98  
Unit: mg/dl  
Mean: 111.00  
SD: 121.00  
Test/Date: 2022/09/16 16:14:38

QC Name : 10C  
Lot No. : 01  
Test Name: URICACI  
Abs. : 0.3105  
Result: 5.15  
Unit: mg/dl  
Mean: 4.90  
SD: 5.40  
Test/Date: 2022/09/16 17:15:35

QC Name : 10C  
Lot No. : 01  
Test Name: TBIL  
Abs. : 0.1091  
Result: 1.82  
Unit: mg/dl  
Mean: 1.70  
SD: 2.00  
Test/Date: 2022/09/16 17:23:26

QC Name : 10C  
Lot No. : 01  
Test Name: ALBUMIN  
Abs. : 0.4569  
Result: 3.50  
Unit: mg/dl  
Mean: 3.20  
SD: 3.80  
Test/Date: 2022/09/16 17:32:49

QC Name : 10C  
Lot No. : 01  
Test Name: TOTPRO  
Abs. : 0.3285  
Result: 4.63  
Unit: mg/dl  
Mean: 4.90  
SD: 5.60  
Test/Date: 2022/09/16 17:41:32

QC Name : 10C  
Lot No. : 01  
Test Name: UREA  
Abs. : 0.2320  
Result: 34.38  
Unit: mg/dl  
Mean: 29.00  
SD: 35.00  
Test/Date: 2022/09/16 18:03:57

QC Name : 10C  
Lot No. : 01  
Test Name: CREAT  
Abs. : 0.0181  
Result: 1.45  
Unit: mg/dl  
Mean: 1.40  
SD: 1.60  
Test/Date: 2022/09/16 18:10:55