



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

**EQAP CODE No. :** 3837

**Distribution No.:** 157-J

**Month/Year:** September/2022

**Instrument ID:** SYSMEX XN350 (14501)

**Name & Contact No. of PT Co-ordinator:** Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

**Date of issue & status of the report:** 27-09-2022[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	6.95	6.78	13.73	12.25	0.0000	0.93	0.17	0.1	0.0000	0.63
RBC x10 <sup>6</sup> /µl	1	4.63	4.6	9.23	9.01	0.0000	0.87	0.03	0.03	0.0000	0.00
Hb g/dl	1	13.4	13.4	26.8	27	0.0000	-0.27	0	0.1	0.0000	-1.35
HCT%	1	41.4	41.2	82.6	83	0.0000	-0.10	0.2	0.4	0.0000	-0.45
MCV-fl	1	89.6	89.4	179	183.9	0.0000	-0.71	0.2	0.4	0.0000	-0.30
MCH-Pg	1	29.1	28.9	58	60.1	0.0000	-1.23	0.2	0.2	0.0000	0.00
MCHC-g/dl	1	32.5	32.4	64.9	65	0.0000	-0.03	0.1	0.3	0.0000	-0.54
Plt. x10 <sup>3</sup> /µl	1	196	181	377	397	0.00	-0.56	15	6	0.00	1.73
Retic %	2	2.8	2.5	5.3	6.5	0.00	-0.29	0.3	0.25	0.00	0.22

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT
<b>DLC%</b>	3	Nrbcs=1 , Poly=54 L=4, E=0, Mono/Promono=0 , B1=3 P.M.=4, Mye=22, Meta=11, Other=	
<b>RBC Morphology</b>	3	moderate anisocytosismicrocytic with mild hypochromic appearance	
<b>Diagnosis</b>	3	Chronic myeloproliferative disorder	

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 157--J	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	101	101	70.3	65.35	2.97	2.97	26.73	31.68
<b>RBC x10<sup>6</sup>/µl</b>	1	101	101	68.32	66.34	0.99	2.97	30.69	30.69
<b>Hb g/dl</b>	1	101	101	66.34	69.31	2.97	1.98	30.69	28.71
<b>HCT%</b>	1	101	101	62.38	68.32	8.91	3.96	28.71	27.72
<b>MCV-fl</b>	1	101	101	63.37	69.31	6.93	4.95	29.7	25.74
<b>MCH-Pg</b>	1	101	101	62.38	62.38	4.95	5.94	32.67	31.68
<b>MCHC-g/dl</b>	1	101	101	64.36	65.35	6.93	5.94	28.71	28.71
<b>Plt. x10<sup>3</sup>/µl</b>	1	101	101	56.44	67.33	7.92	3.96	35.64	28.71
<b>ReticCount%</b>	2	101	62	98.39	85.48	1.61	3.23	0	11.29
<b>PS Assessment</b>	3	101	87	Satisfactory ;, Borderline Sat. :, Unsatisfactory :					

**\*Comments:**

**1). Among Lab (EQA) : Results acceptable.**

**2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----