

VIEW LAB MONTHLY SUMMARY

Lab Name JAYDEE DIAGNOSTIC LABORATORY
 Lab No 4890
 Month August
 Year 2022
 Constituent Group Chemistry I

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/08/2022

Date of Report Published : 05/09/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Biosystems	329	186.51	6.04	11.27	181 mg/dL	-0.49	1.24
2	UREA	Urease UV / GLDH	Biosystems	301	58.79	8.14	4.79	66.9 mg/dL	1.69	0.55
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Biosystems	228	2.55	9.84	0.25	2.8 mg/dL	1.00	0.03
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Biosystems	276	3.62	11.20	0.41	4.1 mg/dL	1.18	0.05
5	T-PROTEIN	Biuret - Colorimetric	Biosystems	313	6.01	7.27	0.44	5.3 g/dL	-1.62	0.05
6	ALBUMIN	BCG - colorimetric	Biosystems	301	3.57	7.32	0.26	3 g/dL	-2.18	0.03
7	CALCIUM	Arsenazo III	Biosystems	215	11.48	8.22	0.94	9.4 mg/dL	-2.20	0.13
8	URIC ACID	Enzymatic / Uricase Colorimetric	Biosystems	296	5.88	8.11	0.48	4.5 mg/dL	-2.89	0.06
9	CHOLESTEROL	CHOD-PAP	Biosystems	307	121.55	5.84	7.10	124 mg/dL	0.35	0.81
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	277	102.75	8.78	9.02	90 mg/dL	-1.41	1.08
11	HDL	Direct method / Enzymatic colorimetric	Biosystems	210	28.70	14.21	4.08	29 mg/dL	0.07	0.56
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1027	139.51	3.22	4.49	146 mmol/L	1.45	0.28
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1065	4.59	5.63	0.26	4.84 mmol/L	0.97	0.02
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	869	102.65	4.14	4.25	105.3 mmol/L	0.62	0.29
15	AST	UV kinetic(with & without PLP (P-5-P))	Biosystems	276	74.61	12.03	8.97	83 U/L	0.93	1.08
16	ALT	UV kinetic(with & without PLP (P-5-P))	Biosystems	238	53.10	11.40	6.05	54 U/L	0.15	0.78