

VIEW LAB MONTHLY SUMMARY

Lab Name	JAYDEE DIAGNOSTIC LABORATORY	Details About Robust Analysis								
Lab No	4890	Detail About Monthly Summary								
Month	June	Detail about SDI								
Year	2022									
Constituent Group	Chemistry I									
		Click on the analyte to view Graphical Data			All Analyser Result		Print			
		Date of Result Entered : 18/06/2022								
		Date of Report Published : 07/07/2022								
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV SD		Your Value	SDI	U
1	GLUCOSE	GOD-POD	Biosystems	305	140.31	5.26	7.38	146 mg/dL	0.77	0.85
2	UREA	Urease UV / GLDH	Biosystems	298	81.35	7.74	6.30	85.2 mg/dL	0.61	0.73
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Biosystems	234	7.59	8.07	0.61	8 mg/dL	0.67	0.08
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Biosystems	288	5.66	9.30	0.53	5.6 mg/dL	-0.11	0.06
5	T-PROTEIN	Biuret - Colorimetric	Biosystems	314	4.50	7.51	0.34	4.3 g/dL	-0.59	0.04
6	ALBUMIN	BCG - colorimetric	Biosystems	299	2.73	9.36	0.26	2.6 g/dL	-0.51	0.03
7	CALCIUM	Arsenazo III	Biosystems	201	8.05	9.40	0.76	9.7 mg/dL	2.18	0.11
8	URIC ACID	Enzymatic / Uricase Colorimetric	Biosystems	290	10.22	6.55	0.67	8.1 mg/dL	-3.17	0.08
9	CHOLESTEROL	CHOD-PAP	Biosystems	287	87.53	7.08	6.19	94 mg/dL	1.04	0.73
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	272	145.75	7.23	10.54	168 mg/dL	2.11	1.28
11	HDL	Direct method / Enzymatic colorimetric	Biosystems	216	22.51	15.27	3.44	24 mg/dL	0.43	0.47
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	961	113.20	5.02	5.68	118 mmol/L	0.85	0.37
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	972	2.04	11.50	0.24	2.29 mmol/L	1.06	0.02
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	796	82.39	7.07	5.82	88.6 mmol/L	1.07	0.41
15	AST	UV kinetic(with & without PLP (P-5-P))	Biosystems	263	55.24	8.97	4.95	53 U/L	-0.45	0.61
16	ALT	UV kinetic(with & without PLP (P-5-P))	Biosystems	255	104.71	12.06	12.63	64 U/L	-3.22	1.58

17	ALP	PNP AMP kinetic	Biosystems	255	370.81	10.77	39.94	173 U/L	-4.95	5.00
18	AMYLASE	CNPG3	Biosystems	104	36.09	13.25	4.78	41 U/L	1.03	0.94
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out