

VIEW LAB MONTHLY SUMMARY

Lab Name	JAYDEE DIAGNOSTIC LABORATORY	Details About Robust Analysis								
Lab No	4890	Detail About Monthly Summary								
Month	May	Detail about SDI								
Year	2022									
Constituent Group	Chemistry I									
Click on the analyte to view Graphical Data						All Analyser Result		Print		
Date of Result Entered : 17/05/2022										
Date of Report Published : 08/06/2022										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV SD		Your Value	SDI	U
1	GLUCOSE	GOD-POD	Biosystems	319	184.05	7.02	12.92	168 mg/dL	-1.24	1.45
2	UREA	Urease UV / GLDH	Biosystems	311	58.00	7.74	4.49	53.3 mg/dL	-1.05	0.51
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Biosystems	243	2.53	9.93	0.25	2.6 mg/dL	0.28	0.03
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Biosystems	267	3.62	10.03	0.36	5.5 mg/dL	5.22	0.04
5	T-PROTEIN	Biuret - Colorimetric	Biosystems	302	5.85	5.28	0.31	6.7 g/dL	2.75	0.04
6	ALBUMIN	BCG - colorimetric	Biosystems	301	3.54	6.64	0.24	3.7 g/dL	0.68	0.03
7	CALCIUM	Arsenazo III	Biosystems	197	11.49	8.78	1.01	11.9 mg/dL	0.41	0.14
8	URIC ACID	Enzymatic / Uricase Colorimetric	Biosystems	269	5.77	6.62	0.38	5.6 mg/dL	-0.44	0.05
9	CHOLESTEROL	CHOD-PAP	Biosystems	305	120.46	6.46	7.78	132 mg/dL	1.48	0.89
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	297	102.67	8.16	8.38	106 mg/dL	0.40	0.97
11	HDL	Direct method / Enzymatic colorimetric	Biosystems	198	27.91	10.39	2.90	32 mg/dL	1.41	0.41
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1024	139.25	3.38	4.71	145.4 mmol/L	1.31	0.29
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1002	4.58	5.70	0.26	4.77 mmol/L	0.73	0.02
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	785	102.75	4.07	4.18	99.3 mmol/L	-0.82	0.30
15	AST	UV kinetic(with & without PLP (P-5-P))	Biosystems	278	74.19	10.04	7.44	77 U/L	0.38	0.89
16	ALT	UV kinetic(with & without PLP (P-5-P))	Biosystems	252	52.06	13.02	6.78	42 U/L	-1.48	0.85

17	ALP	PNP AMP kinetic	Biosystems	284	104.93	11.51	12.08	209 U/L	8.62	1.43
18	AMYLASE	CNPG3	Biosystems	96	45.67	14.18	6.48	74 U/L	4.37	1.32
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out