

VIEW LAB MONTHLY SUMMARY

Lab Name	JAYDEE DIAGNOSTIC LABORATORY	Details About Robust Analysis								
Lab No	4890	Detail About Monthly Summary								
Month	April	Detail about SDI								
Year	2022									
Constituent Group	Chemistry I									
		Click on the analyte to view Graphical Data			All Analyser Result		Print			
		Date of Result Entered : 19/04/2022								
		Date of Report Published : 05/05/2022								
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV SD		Your Value	SDI	U
1	GLUCOSE	GOD-POD	Biosystems	304	222.05	6.04	13.40	225 mg/dL	0.22	1.54
2	UREA	Urease UV / GLDH	Biosystems	285	34.84	10.00	3.48	34.5 mg/dL	-0.10	0.41
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Biosystems	233	4.87	8.96	0.44	4.1 mg/dL	-1.77	0.06
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Biosystems	262	2.89	11.96	0.35	3 mg/dL	0.32	0.04
5	T-PROTEIN	Biuret - Colorimetric	Biosystems	292	5.48	6.39	0.35	5 g/dL	-1.37	0.04
6	ALBUMIN	BCG - colorimetric	Biosystems	287	3.36	8.93	0.30	2.9 g/dL	-1.53	0.04
7	CALCIUM	Arsenazo III	Biosystems	192	10.59	8.42	0.89	9.8 mg/dL	-0.89	0.13
8	URIC ACID	Enzymatic / Uricase Colorimetric	Biosystems	283	7.26	7.71	0.56	6 mg/dL	-2.25	0.07
9	CHOLESTEROL	CHOD-PAP	Biosystems	296	112.92	7.47	8.44	114 mg/dL	0.13	0.98
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	270	125.53	8.58	10.77	135 mg/dL	0.88	1.31
11	HDL	Direct method / Enzymatic colorimetric	Biosystems	199	28.57	13.60	3.89	32 mg/dL	0.88	0.55
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	986	135.62	2.99	4.05	138.9 mmol/L	0.81	0.26
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	958	3.94	5.03	0.20	4.06 mmol/L	0.61	0.01
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	739	105.68	3.86	4.08	102.4 mmol/L	-0.80	0.30
15	AST	UV kinetic(with & without PLP (P-5-P))	Biosystems	253	124.90	10.28	12.83	128 U/L	0.24	1.61
16	ALT	UV kinetic(with & without PLP (P-5-P))	Biosystems	262	59.93	13.47	8.07	50 U/L	-1.23	1.00

17	ALP	PNP AMP kinetic	Biosystems	287	131.26	13.37	17.55	196 U/L	3.69	2.07
18	AMYLASE	CNPG3	Biosystems	96	79.35	13.77	10.93	74 U/L	-0.49	2.23
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out