

VIEW LAB MONTHLY SUMMARY

Lab Name	JAYDEE DIAGNOSTIC LABORATORY	Details About Robust Analysis								
Lab No	4890	Detail About Monthly Summary								
Month	March	Detail about SDI								
Year	2022									
Constituent Group	Chemistry I									
		Click on the analyte to view Graphical Data			All Analyser Result		Print			
		Date of Result Entered : 19/03/2022								
		Date of Report Published : 06/04/2022								
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE	GOD-POD	Biosystems	289	108.93	5.90	6.43	108 mg/dL	-0.14	0.76
2	UREA	Urease UV / GLDH	Biosystems	272	78.40	7.49	5.88	80 mg/dL	0.27	0.71
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Biosystems	239	3.99	10.72	0.43	3.8 mg/dL	-0.44	0.06
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Biosystems	259	1.64	18.26	0.30	1.7 mg/dL	0.20	0.04
5	T-PROTEIN	Biuret - Colorimetric	Biosystems	304	5.03	7.07	0.36	5.1 g/dL	0.20	0.04
6	ALBUMIN	BCG - colorimetric	Biosystems	275	3.11	7.94	0.25	2.9 g/dL	-0.85	0.03
7	CALCIUM	Arsenazo III	Biosystems	198	8.78	8.68	0.76	9.8 mg/dL	1.34	0.11
8	URIC ACID	Enzymatic / Uricase Colorimetric	Biosystems	290	8.37	7.91	0.66	5.2 mg/dL	-4.79	0.08
9	CHOLESTEROL	CHOD-PAP	Biosystems	284	106.70	6.91	7.38	106 mg/dL	-0.09	0.88
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	257	204.33	6.85	13.99	230 mg/dL	1.84	1.75
11	HDL	Direct method / Enzymatic colorimetric	Biosystems	195	26.10	12.68	3.31	28 mg/dL	0.57	0.47
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	920	124.76	4.58	5.71	131.4 mmol/L	1.16	0.38
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	882	2.97	6.60	0.20	3.18 mmol/L	1.07	0.01
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	702	91.37	5.08	4.64	90.6 mmol/L	-0.17	0.35
15	AST	UV kinetic(with & without PLP (P-5-P))	Biosystems	249	254.10	11.03	28.02	180 U/L	-2.64	3.55
16	ALT	UV kinetic(with & without PLP (P-5-P))	Biosystems	240	163.21	13.00	21.22	210 U/L	2.20	2.74

17	ALP	PNP AMP kinetic	Biosystems	263	202.77	11.79	23.91	212 U/L	0.39	2.95
18	AMYLASE	CNPG3	Biosystems	96	71.58	12.81	9.17	65 U/L	-0.72	1.87
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out