



CC-2725

**CERTIFICATE OF CALIBRATION**

Certificate No: CC-22-23-L-461/01

Page. No : 1 of 1

Customer Name & Address M/s. VINAYAK DIAGNOSTIC CENTRE No: 7, Jambulingam Street, Nungambakkam, Chennai - 600 034.	ULR No.	CC272522000007477F
	SRF No.	VCI-22-23-L-461
	Date of Receipt	24.09.2022
	Date of Calibration	24.09.2022
	Due Date for Calibration	23.09.2023
	Date of Issue	26.09.2022

Details of Unit Under Calibration

Description	Thermohygrometer	Make / Model	HTC/HTC-1
Range	10 to 99%RH, -50°C to 70°C	Operating Range	20 to 90%RH, 10 to 50°C
Resolution	1%RH, 0.1°C	Condition of UUC	Good
Serial Number	--	Instrument Location	--
ID Number	DTH-01	Cal. At	Thermal Lab

Environmental Condition		Calibration Method Used	
Temperature (°C)	25±4	Cal Procedure No	VCI-L03-WI-THER-008
Humidity (%RH)	30 - 75		

Standard Used

Sl. No.	Description	ID No/ SL .No	Certificate Number	Make & Model	Traceability	Validity
1	Digital Temperature Humidity Indicator with Sensor	VCI-T-DTHI-038	TSC/22-23/3002-2	Tempser/TDPC-9696	National Standard	05.06.2023

Results of Calibration**Humidity:@25°C**

Sl. No.	Standard Reading (%RH)	UUC Reading (%RH)	Error (%RH)	Measurement Uncertainty ± (%RH)
01	20.63	20	-0.63	2.24
02	50.81	50	-0.81	2.24
03	90.95	90	-0.95	2.24

Temperature: @ 50% RH

Sl. No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Measurement Uncertainty ± (°C)
01	10.16	10.0	-0.16	0.37
02	25.23	25.1	-0.13	0.37
03	50.37	50.1	-0.27	0.37

Remarks:

- UUC is Defined as the Unit Under Calibration.
- This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- The reported results are valid at the time of under stated condition of measurement.
- Calibration of the UUC are traceable to National / International Standards.
- The recalibration interval shall be decided by the user.
- Nabl-133 guidelines are adopted for use of NABL symbol.
- The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor k=2.

Checked & Issued by:

M.Manju

M.Manju

(Engineer - Calibration)



Authorised by:

C.Dhinakaran

(Technical / Quality Manager)

No.16, 6th Street, Annai Sathya Nagar, Poothapedu Ramapuram, Chennai - 89.

Ph: +91 44 2476 0533 /0633, Mob: 90031 20649, 9789893643, 8807402558

e: venuscallab@gmail.com, info@venuscal.com | www.venuscal.com