



Calibration Certificate

Cell : 99804 08580

Maruthi Enterprises

103, 4th Main, Gangappa Block, Ganganagar, Bangalore - 560 032.
E-mail : maruthi.enterprises@yahoo.com

CALIBRATION CERTIFICATE

CUSTOMER:- REHA DIAGNOSTIC LABORATORY,
#15/A, OLD POLICE STATION ROAD,
KONANAKUNTE,
BANGALORE-62.

CERTIFICATE NO :- MEC - 14523
CALIBRATION DATE :- 01/01/2022

DESCRIPTION OF ITEM :- BLOOD MIXER

RANGE:- 0-180 RPM

MODEL :- BLOD MIXER -01

S.NO :- B.MIXER -01

DETAILS OF EQUIPMENT USED FOR CALIBRATION :-

S.NO	DESCRIPTION	CERTIFIED BY	CERTIFICATE NO	VALIDITY
1	DIGITAL NON CONTACT TACHOMETER	BCL	SRF NO:- 4375 -01	21/02/2022

CALIBRATION DATA :-

S.NO	CALIBRATION POINTS(IN RPM)	DISPLAY READING	OBSERVED ERROR	REMARKS
1	30 RPM	32 RPM	2 RPM	OK
2	60 RPM	62 RPM	2 RPM	OK
3	120 RPM	123 RPM	1 RPM	OK
4	180 RPM	179 RPM	1 RPM	OK

T CALIBRTION BUE ON:- 01/01/2023

NOTE:-

1. THIS CALIBRTION CERTIFICATE RELATES ONLY TO THE ABOVE DUC.
2. CALIBRTION CERTIFICATE SHALL NOT BE REPRODUCED EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVED FROM BCL.
3. RESULT REPORTED ARE VALID AT THE TIME OF AND UNDER THE STATED CONDITONOF MEASUREMENT.
4. ANY ERROR IN THIS REPORT SHOULD BD BROUGHT TO OUR KNOWLEDGEWITHIN 45 DAYS FROM THE DATE OF THIS CERTIFICATE.

CALIBRATED BY :-
MURTHY B.



CHECKED BY:-
DRISHYA K.K.