



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2022



PC-1024

Lab Name Dr.KHANNA PATHCARE PVT.LTD Lab No 5284
 Constituent Group Chemistry II Date of Result Entered : 17/08/2022
 PT item Lyophilized Serum Date of Report Published : 05/09/2022

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1248	109.27	9.28	10.14	121 mg/dL	1.16	0.57
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	727	24.41	15.14	3.70	19.5 mg/dL	-1.33	0.27
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	663	6.21	13.54	0.84	5.7 mg/dL	-0.61	0.07
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1041	4.67	13.69	0.64	4.5 mg/dL	-0.27	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	980	5.90	10.31	0.61	5.2 g/dL	-1.15	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	905	3.45	10.31	0.36	3.6 g/dL	0.42	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	795	4.27	18.00	0.77	2.7 mg/dL	-2.04	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1091	117.37	10.60	12.44	113.5 mg/dL	-0.31	0.75
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1034	189.36	11.26	21.32	187 mg/dL	-0.11	1.33

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Dr.KHANNA PATHCARE PVT.LTD

C/O SANJEEVAN MEDICAL RESEARCH CENTRE Pvt Ltd, 4826/24, ANSARI ROAD, DARYAGANJ DELHI

DELHI110002

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****