



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments  
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092  
 Phone / Fax : (011) 22453259, 22434362, 9599482307  
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL LABORATORY CENTRIFUGE

Certificate No:	22900004090	Issued On:	05/09/2022
ULR No:	CC214622900004090F	Calibration Date:	02/09/2022
Job Identification No:	ICS/C/AS/09/4090	Next Calibration Date:	01/09/2023
Ref. No:	SRF- Dated: 30/08/2022		

#### CALIBRATED FOR:

Dr. Khanna's Pathcare Pvt. Ltd.  
 C/o Sanjeevan Medical Research Centre Pvt. Ltd.  
 24, Ansari Road,  
 Daryaganj, New Delhi-110002.

#### EQUIPMENT DESCRIPTION

Name	RPM Measurement for Digital Laboratory Centrifuge		
Make/Model	Remi / R-4C	Visual Inspection	Ok
Serial No.	ZGAN-02247	I.D. No.	SMRC/GEN/LAB/CEN-01
Range	Up to 4000 RPM	Least Count	10 RPM
Location	---	Calibration Site	On-Site

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

#### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

#### USED EQUIPMENT DETAILS

Not Applicable

#### PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-RPM01  
(Measurement Method)

#### RESULTS: Mechanical Calibration MEASUREMENT: (in RPM)

CENTRIFUGE' READING	CORRECTED VALUE OF TACHOMETER READING (AVERAGE)	UNCERTAINTY AT APPROX 95% C.L AND COVERAGE FACTOR K=2
1140	1142.4	± 6
1470	1474.6	± 6
1890	1897.5	± 7
2510	2520.5	± 7
3320	3333.2	± 8

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
 3. Total liability of our Organisation is limited to the invoiced amount.  
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.  
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

RAJESH KAPOOR  
Operation-Head