

SHAN

SHAN BIOTECH & DIAGNOSTICS

"SHAN TOWERS" No. 8, Baskara Street,
Rangarajapuram, Kodambakkam, Chennai - 600 024.
Phone : 044 - 2481 6330, 2480 4186
E-mail : msk_1804@rediffmail.com
shanbiotechsiva@gmail.com

Calibration Certificate

Model Name: PRIESTEST TOUCH PLUS Sr. No: ATP4710319RBK Malar Diagnostics No: P/131, Pillayar Koil Sttreet, Kelambakkam-603 103	Certificate No. : Analyser-ATP-004 / 2022-2023
	Date of Calibration : 13-09-2022
	Calibration Due on : 12-09-2023

STANDARD USED FOR CALIBRATION






Parameters	Inst. Name & Sr. No.	Cert. No.	Calibrated by	Validity
Temperature Calibration	Digital Thermometer	UCS/CAL/21100315	Universal Instrument	31/10/2022
Temperature Calibration	Digital Thermometer	UCS/CAL/21100314	Universal Instrument	31/10/2022
Film readings Calibration	Spectrophotometer, Make:Shimandzu, Model No.:UV-1800, Sr. No. A-1145 4702871	REG/2S/2207	QC dept. of Reagent Robonik india pvt.ltd	20/07/2022

OBSERVATIONS (Temperature Calibration)

Temperature Details	Incubator Temperature	Cuvette Temperature
Thermometer Sr. No.	025	012
Obtained Value	37.5°C	37.4°C
Range	37°C (± 1°C)	37°C (± 0.5°C)

OPTICAL FILM READING CALIBRATION

Wavelength	Film type	Range	Absorbance Obtained
340 nm	1SP	0.549 - 0.619	0.590
	1HP + 1SP	0.842 - 0.975	0.914
	1TP	1.627 - 1.853	1.737
	2DP + 1HP	2.643 - 3.427	2.714
405 nm	1SP	0.307 - 0.348	0.318
	1TP	0.929 - 1.062	1.007
	2DP + 1HP	1.432 - 1.642	1.572
	3TP	2.809 - 3.318	3.034
510 nm	1SP	0.299 - 0.337	0.300
	1TP	0.913 - 1.054	0.997
	2DP + 1HP	1.386 - 1.576	1.505
	3TP	2.784 - 3.232	2.997
545 nm	1SP	0.264 - 0.313	0.286
	1TP	0.844 - 0.956	0.918
	2DP + 1HP	1.287 - 1.446	1.393
	3TP	2.512 - 2.849	2.790
578 nm	1SP	0.265 - 0.302	0.292
	1TP	0.791 - 0.904	0.874
	2DP + 1HP	1.200 - 1.391	1.337
	3TP	2.387 - 2.702	2.660
630 nm	1SP	0.300 - 0.352	0.331
	1TP	0.977 - 1.115	1.065
	2DP + 1HP	1.443 - 1.648	1.584
	3TP	2.806 - 3.210	3.165

Biochemical Analyser bearing serial no. ATP4710319RBK above parameters and found to be in range criteria. Calibrated By:

GSTIN : 33AUGPS3698Q1ZM

DL No. 3960/MZII/20 B & 4125/MZII/21 B DL 27-06-2018