



CERTIFICATE OF CALIBRATION

Certificate No: CC-22-23-L-441/01

Page No: 1 of 1

Customer Name & Address

M/s. MALAR DIAGNOSTICS
NO. 3/129, A Pillayar Koll Street,
K.R.R Complex,
Kelambakkam - 603 103.

ULR No.	CC27252200007056F
SRF No.	VCI-22-23-L-441
Date of Receipt	12.09.2022
Date of Calibration	12.09.2022
Due Date for Calibration	11.09.2023
Date of Issue	13.09.2022

Details of Unit Under Calibration

Description	Micro Pipette	Make	-
Range	10µl	Model	-
Resolution	NA	Material	pvc
Serial Number	NC452873	Operating Range	10µl
ID Number	-	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	-

Environmental Condition				Calibration Method Used	
Temperature (°C)	23±2	Humidity (%RH)	55±15	National / International Standard	ISO 8655-6:2002, ISO 8655-2:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.4	Cal Procedure No	VCI-L03-WI-MASS-003

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	VCI-M-WB-017 / 55492	CC-21-22-S-994/01	Wensar / MAB 250	National Standards	14.10.2022

Z Factor: 1.00319

Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
		10.05	10.09	10.15	10.19	10.23					
1	10	10.05	10.09	10.15	10.19	10.23	10.15	0.15	0.07	0.20	
		10.07	10.13	10.17	10.21	10.25					

Note: Conformity Statement

Pass:

Fail:

Remarks

1. UUC is Defined as the Unit Under Calibration.
2. The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor k=2.
3. This Calibration Certificate relates only to the above UUC & Reported results are valid at the time of and under the stated conditions of measurements.
4. Calibration of the UUC are traceable to National / International Standards.
5. This report shall not be reproduced in full/ part without prior permission of Venus Calibration and Instruments.
6. The recalibration Interval shall be determined on the user requirements.
7. Nabl-133 guidelines are adopted for use of NABL symbol.
8. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.
9. Our Nabl Scope is Upto 10µl to 10ml.
10. The Deviation is within MPE, So after calibration is not required.

Checked & Issued by:

M. Manju
M.Manju

(Engineer - Calibration)



Authorised by:

C.Dhinakaran

(Technical/Quality Manager)

No.16, 6th Street, Annai Sathya Nagar, Poothapedu Ramapuram, Chennai - 89.

Ph: +91 44 2476 0533 /0633, Mob: 90031 20649, 9789893643, 8807402558

e: venuscallab@gmail.com, info@venuscal.com | www.venuscal.com