

MEDICAL TRUST HOSPITAL (680)
Outlier And Analyte Summary Report

Outlier Details For Cycle No 220207 and Sample No 02

Report Date : 31/08/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Creatinine	VITROS 5600 (Dry Slide)	0.80	mg/dl	2.94
<p>Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated 🕒 : Delayed Result Entry * : Not considered for evaluation.</p>				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

MEDICAL TRUST HOSPITAL (680)

Outlier And Analyte Summary Report

Result Details For Cycle No 220207 and Sample No 02


Report Date : 31/08/2022

Instrument : VITROS 5600 (Dry Slide)

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	66.00	--	-1.89	-1.05
@ Urea	mg/dl	20.00	--	0.43	0.69
! Creatinine	mg/dl	0.80	--	2.94	0.95
@ Bilirubin Total	mg/dl	0.10	--	0.00	-0.40
@ Total Protein	g/dL	5.70	--	0.00	0.33
@ Albumin	g/dL	3.60	--	-0.60	-0.03
@ Uric Acid	mg/dl	4.00	--	-0.18	-0.68
@ Calcium	mg/dl	7.60	--	1.66	1.60
@ Phosphorous Inorganic	mg/dl	3.70	--	1.36	0.84
@ Magnesium	mg/dl	1.50	--	0.00	-0.23
@ Cholesterol	mg/dl	132.00	--	0.67	1.48
@ Triglycerides	mg/dl	135.00	--	1.15	0.59
@ HDL	mg/dl	29.00	--	0.00	0.35
@ LDL	mg/dl	78.00	--	0.08	-0.25
@ ALT	U/L	6.00	--	-1.01	-1.47
@ AST	U/L	19.00	--	0.00	-0.18
@ ALP	U/L	55.00	--	0.52	0.20
@ CK	U/L	34.00	--	0.20	-0.17
@ GGT	U/L	26.00	--	0.41	0.54
@ LDH	U/L	105.00	--	0.42	0.06
@ Amylase	U/L	61.00	--	-0.06	-0.70
@ Chloride	mmol/L	84.00	--	-1.69	-1.99
@ Lipase	U/L	118.00	--	0.00	11.18
@ Sodium	mmol/L	136.00	--	-0.29	-0.18
@ Potassium	mmol/L	3.80	--	0.48	0.33

Legend @ : z score ≤ 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score ≥ 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	25
Not Evaluated Parameters	0
Evaluated Parameters	25
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070