

MEDICAL TRUST HOSPITAL (680)
Outlier And Analyte Summary Report

Outlier Details For Cycle No 220207 and Sample No 01

Report Date : 31/08/2022

Analyte	Instrument	Result Value	Standard Unit
X Specific Gravity	Labumat 2	1.010	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

MEDICAL TRUST HOSPITAL (680)

Outlier And Analyte Summary Report

Result Details For Cycle No 220207 and Sample No 01


Report Date : 31/08/2022

Instrument : Labumat 2

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6.5	6,6.5,7
X Specific Gravity	-	1.010	1.015,1.020,1.025
@ Glucose (urine)	-	4+	+2,+3,+4
@ Protein (Urine)	-	Negative	+1,Neg,Tr
@ Ketone (urine)	-	Negative	+1,Neg,Tr
@ Blood (urine)	-	1+	+1,Neg,Tr
@ Bilirubin (urine)	-	Negative	+1,Neg
@ Urobilinogen (urine)	-	Normal	+1,Nor
@ Nitrite (urine)	-	Absent	Abs

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Clinical Path - Urine Routine Chemistry	88.89 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report