

**COTEX LAXMI HEALTHCARE PVT. LTD.**

Sawangi Meghe ,Wardha-442001

CALIBRATION CERTIFICATE

1.Customer																															
Name and Address of Customer: SHALINITAI MEGHE HOSPITAL AND RESEARCH CENTER,WANADONGARI,HINGNA NAGPUR,441110	Procedure No. CLH/CAL/20/AWB/SOP/07 Date of Calibration: 21-04-2022 Due Date of Calibration: 20-04-2023 Location of Calibration: IN LAB Equipment Status: OK																														
Ambient Temp.: (25±2) ^o c RH: 45 %RH to 75 %RH																															
2.Description of item																															
Name: MICROPIPETTE Make: NA Model: NA Sr.no. NL484283 Range 0.5-10 µl Specified accuracy: NA	Department: PATHOLOGY LAB Location: PATHOLOGY LAB Asset No.: 3129 Resolution: Calibration Certificate no.: CLH/CAL/SMHRC/CCL/PATH/02 Type: VARIABLE																														
3.Equipment used for Calibration (Master)																															
Name: ANALYTICAL WEIGHT BALANCE Make/ Model: CONTECH/CAI35 Sr.No. 3000113 Certificate no. WM-01 Validity 09-05-2023																															
The Masters are traceable to National/International Standards.																															
4.(1) Visual Test Inspection																															
<table border="1"><thead><tr><th>SR.NO.</th><th>TEST</th><th>REMARK</th></tr></thead><tbody><tr><td>1</td><td>Device clean and Decontaminated</td><td>OK</td></tr><tr><td>2</td><td>Physical Damage</td><td>OK</td></tr><tr><td>3</td><td>Equipment type(B,BF,CF)</td><td>B</td></tr></tbody></table>	SR.NO.	TEST	REMARK	1	Device clean and Decontaminated	OK	2	Physical Damage	OK	3	Equipment type(B,BF,CF)	B																			
SR.NO.	TEST	REMARK																													
1	Device clean and Decontaminated	OK																													
2	Physical Damage	OK																													
3	Equipment type(B,BF,CF)	B																													
B-Body Type, BF-Body Float Type, CF-Cardiac Float Type.																															
5. Performance Test																															
<table border="1"><thead><tr><th>SR.NO.</th><th>CALIBRATION POINT (µl)</th><th>NOMINED VALUE OF VOLUME (µl)</th><th>OBSERVED VALUE VOLUME ON MASTER (µl)</th><th>ERROR IN µl</th></tr></thead><tbody><tr><td>1</td><td>0.5</td><td>0</td><td>0</td><td>0</td></tr><tr><td>2</td><td>2</td><td>2</td><td>1.9991</td><td>0.0009</td></tr><tr><td>3</td><td>6</td><td>6</td><td>6</td><td>0</td></tr><tr><td>4</td><td>8</td><td>8</td><td>8</td><td>0</td></tr><tr><td>5</td><td>10</td><td>10</td><td>10</td><td>0</td></tr></tbody></table>	SR.NO.	CALIBRATION POINT (µl)	NOMINED VALUE OF VOLUME (µl)	OBSERVED VALUE VOLUME ON MASTER (µl)	ERROR IN µl	1	0.5	0	0	0	2	2	2	1.9991	0.0009	3	6	6	6	0	4	8	8	8	0	5	10	10	10	0	
SR.NO.	CALIBRATION POINT (µl)	NOMINED VALUE OF VOLUME (µl)	OBSERVED VALUE VOLUME ON MASTER (µl)	ERROR IN µl																											
1	0.5	0	0	0																											
2	2	2	1.9991	0.0009																											
3	6	6	6	0																											
4	8	8	8	0																											
5	10	10	10	0																											
Page 1																															

Note: (1) The Reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the Standard uncertainty by the coverage factor $k=2$, which corresponds to coverage probability of approximately 95% for normal distribution.


(2) The certificate refers only to the particular item submitted for calibration UUC stands for unit under calibration.

(3) The calibration result reported in the certificate are valid at the time of and under the stated condition of measurement.

(4) The certificate point were selected as per customer specifications and calibration report traceable to national or international standard.

(5) This certificate shall not be reproduced except in full unless written permission for the publication of an approved abstract has been obtained from the technical manager of COTEX LAXMI HEALTHCARE PVT. LTD. WARDHA.


Calibrated By



Calibration Engineer



Approved By



Technical Manager

National Accreditation Board for Testing and Calibration Laboratories

Doc No.: 24 Specific Criteria for calibration of Medical Devices

Issue No.:01 Issue Date:01/01/2021 Amend No.:01 Amend Date:01/01/2021 Page 2