

M31112041N

Quality Product and Service for Medical Laboratory & Biotech Instruments, Consumable & Reagents "An ISO 9001:2015 Certified Company" SAGAR GALAXY, IInd Floor, LSC, Mayur Vihar - II, New Delhi -110091 Tel.: 011-41004431,

E-mail: priman@primanindia.com / email: admin@primanindia.com Website: www.primanindia.com

Date: 8th Sep 2022

CALIBRATION CERTIFICATE

This is to certify that Instrument model Fully Automated Biochemistry Analyzer Model BS-240 (Sr. No.YX-8A002065) Installed at Central Diagnostic Laboratory Address: Rajendra Memorial Institute of Medical Sciences, Patna, Bihar - 800007, has been calibrated by MINDRAY Biochemistry Calibrator MULTI SERA CALIBRATOR (Lot No. 150122002 EXP; 31-12-2023), During Calibration, Instrument connected through UPS.

> Lamp Voltage: 12 V checked and found 11.98 Gain of all Filters are 48000-60000. As per the Manufacturer's Guideline the gain should be in range of (40,000 - 60,000)

Next Calibration for the Instrument is due after six month from the date of this Certificate.

For Priman Instruments Pvt. Ltd.

Authorized Signatory

Date: 08/Sep/2022

Place: NEW DELHI



Rose Medicare Pvt Ltd 2013, Pratrilla Riace. chavaner Maria Routs Aprile HORKAN

GST: 10AABCR2703B1ZF Drug Licence No: 01/01A/2017 Tel: 0612-230 3103/3182/1922 Email Jain@rosemedicare.com Website www.rosemedicare.com Jain International (A Unit of Rose Medicare Pvt Ltd)

Date of Calibration	Next Calibration Due on	Page No.	Total Page
15 / 01 / 2022	(Recommended) 14 / 01 / 2023	1	1

Certificate No. / 1.0

RMPL / 22-23 / 0454 dated 15-01-2022

Party Ref. (W.O. No.)

Calibrated for 1.1

R.M.R.I.M.S

(Name & Address of Customer)

Agam Kuan. PATNA.

Description & Identification 1.2

Name

Centrifuge

Make

Remi

Model

MEDICO PLUS

SL. No. Inst. ID No. **ZEIN - 31864**

Range

. 10 - 4000

Resolution

10 rpm

Accuracy

Date of Receipt 1.3

15/01/2022

Date of Calibration

15 / 01 / 2022

Valid Up to

14 / 01 / 2023

Environment Condition 1.4

Temperature

 $23.0^{\circ} \text{ C} + / - 3^{\circ} \text{ C}$

Caliborated by

40 % - 60 % Relative Humidity

Calibration Procedure 1.5

SOP / Tachometer

Method Used 1.6

Comparison Method with Respect to

Tachometer values.

Received by

RAV GUPTA)

LABLIFE

SCIENTIFIC INSTRUMENTS REPAIRING SERVICES CENTRE & SALE

Opposite: PMCH, Ashok rajpath Road, Patna-800004

Email : lablife 2000@gmail.com MOB.-9031126729,8968536661

GST NO : 1085MPP7011F122

Service Report cum Certificate

Date: 07.10.2022

This is to certified that Frost free Refrigerator Model No -RT 37K3993UZ/HL/2016, 340 Ltrs, Brand-Samsung under random inspection by me, is working satisfactory and temperature of the said fridge checked at different temp. levels and found satisfactory.

Observation: Refrigerator is working OK.

Service Engineer

Name- SUDARSHAN PANDIT signature-Sydys5h 4N PANdit

Opp. PMCH, Ashok Pajpath Patna-800004