



Quality Product and Service for Medical Laboratory & Biotech Instruments, Consumable & Reagents

"An ISO 9001:2008 Certified Company"

SAGAR GALAXY, IInd Floor, LSC, Mayur Vihar - II, New Delhi -110091

Tel.: 22784431, 22784432,

e-mail: priman@primanindia.com / email: sales@primanindia.com

Website: www.primanindia.com

Date: 8th Sep 2022

CALIBRATION CERTIFICATE

This is to certify that Instrument Electrolyte Analyzer Model PE-21 (Sr. No.1740) Installed at Central Diagnostic Laboratory Address: Rajendra Memorial Institute of Medical Sciences. Patna, Bihar - 800007, has been calibrated on every eight hours During Calibration, Instrument connected through UPS.

- ✓ Open & Cleaned the analyzer.
- ✓ Lubricated all the moving and rotating parts.
- ✓ Checked the working of the machine.
- ✓ Ran some samples, they are good.

Also run the control with below mentioned details. The reference values of control are managed by the traceability system in PRIMAN and these are traceable to the International Standards.

Control: Quality Controls

Lot:

2203

Expiry:

09-Sep-2023

Next preventive maintenance is due after six month from the calibration date. (Service report needs to be attached in case of system copy).

Priman Instruments Pvt. Ltd.

New Delhi

Works: D-7 Dev Bhoomi Industrial Estate, Roorkee, Uttarakhand, (U.K.) Chennai: 09884006799, Patna: 08409048542