



ULR NO: CF48JKDG8949 ISSUE DATE: 25.01.2022	CALIBRATION DONE ON : 25.07.2022 CALIBRATION VALID UPTO : 04.07.2023	SRF. & DATE: 3050/935 & 25.01.2022 PAGE : 1 OF 1
CALIBRATED FOR M/S. <b>VEDANTA HOSPITAL AZAMGARH</b>		

**IDENTIFICATION ON UUC (UNIT UNDER CALIBRATION)**

INSTRUMENT NAME	MICRO PIPETTE	SIZE/RANGE	100 -1000 µl
LEAST COUNT	10 µl	CONDITION OF ITEM	GOOD
MANUFACTURE BY	PROLINE	CALIBRATION PERFORMED AT	SITE
SR NO	AB153431	LOCATION	LAB
CALIBRATION ID	ID/LAB/PP-02	REF. STANDARD	ISO-8655-6
ENVIRONMENT CONDITION: TEMPERATURE: 25 ± 5, RELATIVE HUMIDITY: 55 ± 15% RH			

**STANDARD AND MAJOR EQUIPMENTS**

STANDARD & EQUIPMENT	RANGE	CALIBRATING AGENCY	TRACEBILITY CERT. NO.	CERTIFICATION DATE	VALID UPTO
STANDARD WEIGHT	1 MG to 200 g	Weightronics	SDJGH48937/22	25/01/2022	24/01/2023

**CALIBRATION RESULT AT 27 °C**

NORMAL VALUE UUC (µl)	OBSERVED VALUE (µl)	ERROR (µl)	RELATED UNCERTAINTY ON C.F.(K=2) AT 95% CONFIDENCE LEVEL ±
100	100.49	-0.40	0.38 µl
200	200.56	-0.56	0.38 µl
500	501.83	-0.83	0.38 µl
1000	1003.91	-3.91	0.38 µl

.....END OF REPORT.....



Calibration Engineer's Signature

**Spectro Analytical Labs Limited** S-1, GNEPIP, Surajpur Industrial Area, Phase-V, Kasna, Greater Noida-201308 (India)  
 Phone : +91-120-2341250,2341251 || URL : [www.spectro.in](http://www.spectro.in) || Email: [care@spectro.in](mailto:care@spectro.in)

BIS Approved, ISO 9001:2008 & ISO 14001:2004 & OHSAS 18001:2007 Certified Laboratory  
 Please refer to our Website [www.spectro.in](http://www.spectro.in) for Terms & Condition