

CALIBRATION REPORT

Customer Name : Shree Diagnostic Centre.
Address : Shree Hospital, Ganesh Baug,
Murbad Road, Kalyan
Instrument Model : ST-200 Pro Electrolyte Analyzer
Serial no : ST-000325
Calibration Date : 08.12.2021
Calibration Due Date: 07.12.2022

As per the Standard Operating procedure of ST-200 Pro Electrolyte Analyzer, the instrument has been calibrated.

As per method mentioned into manuals/sop, the Known Value Sample/Control run on the analyzer for five times for the Electrolyte parameters. All obtained values are in the range and CV values were checked. CV values were within a normal range as mentioned in performance certificate. This indicated that an instrument is performing accurately and precisely.

Hence, we certified that ST-200 Pro Electrolyte Analyzer is calibrated.



Makarand Thakur
(Service Engineer)