



CALIBRATION RECORD


INSTRUMENT MODEL	BC-5150	SERIAL NO	SR-11004793
TYPE	HEMATOLOGY ANALYZER-5-part	INSTALLED AT	Reliable Diagnostic Centre Kurukshetra
CALIBRATOR USED	Mindray SC-Cal Plus	Calibrator Lot No	PLUS0395
		Calibrator Expiry	25/08/2022

Enclosed:-

1. Assay Sheet of Mindray SC-Cal Plus for Mindray Hematology (Lot-Plus-0395, Expiry-05/09/2022)

Calibration Done -25/08/2022

Next Calibration Due- 24/08/2023

For Labtech

(Authorised Signatory)





Calibration

Mindray Hematology Model :- BC5150

Lot . No	PLUS0395	Analyzer Sr. no	SR-11004793
Exp:-	05/09/2022	Calibration Date	25/08/2022

	WBC	RBC	HGB	MCV	PLT
Target	7.66	4.27	12.5	93.1	255
1	7.69	4.32	13.8	82.0	241
2	7.42	4.17	13.6	82.9	243
3	7.48	4.28	13.7	83.5	246
4	7.56	4.25	13.7	83.4	248
5	7.59	4.13	13.6	83.7	245
6	7.37	4.26	13.7	82.8	249
7	7.39	4.17	13.6	83.2	239
8	7.51	4.14	13.5	83.0	234
9	7.42	4.21	13.6	83.6	250
10	7.60	4.26	13.7	83.2	240
Mean	7.48	4.21	13.6	83.2	244
CV(%)	1.2	1.4	0.6	0.4	2.2
New Factor(%)	102.38	101.41	92.34	111.83	104.58
Old Factor(%)	98.11	99.18	92.00	108.82	112.64



LABTECH

50/41,OPP.Neerja Modi School
Rajat Path Mansrover
Jaipur-302020
Ph:-0141-2783950

TECHNICAL SERVICE REPORT

TECHNICAL SERVICE REPORT NO. .		DATE: 28/8/2022	
NAME <i>Reliable digiprint</i>	MODEL NO. <i>BC-5150</i>	<input type="checkbox"/> INSTALLATION <input type="checkbox"/> WARRANTY <input type="checkbox"/> CMC <input type="checkbox"/> AMC <input type="checkbox"/> CHARGABLE CALL <input type="checkbox"/> PMVISIT <input type="checkbox"/> APPLICATION SUPPORT OTHER	
	SR.NO <i>SR-11004793</i>		
ADDRESS: <i>Centre Kusupoketran.</i>	CALL DETAIL	WORK CARRIED OUT AT <input checked="" type="checkbox"/> SITE <input type="checkbox"/> SERVICE CENTRE	
	DATE TIME COMPLAINT RECD.		
TEL. NO.	JOB COMPLETED		
NAME OF THE OPERATOR:			
PROBLEM REPORTED: <i>machine calibration</i>			
OBSERVATION:			
ACTION TAKEN: <i>Calibration Done next cal - 24/8/2022</i>			
<input type="checkbox"/> FOLLOWING PARTSHAVE BEEN REPALACED <input type="checkbox"/> FOLLOWING PARTS NEED TO BE REPLACED, PLEASE APPROVE		TO BE FILLED BY CUSTOMER	
NO.	DECRIPATION	QTY	TOTAL
	/		/
TOTAL RS .			
FOLLOW-UP-ACTION:		SEAL	DATE
ENGINEER'S PRODUCT SPECILIST'S SIGNATURE		CUSTOMERS SIGNATURE NAME	
TIME:	NAME <i>[Signature]</i>	BRANCH H.O.	
		RECEIVED NO;	CHECKED BY:
NOTE: Parts replaced are chargeable except during warranty period consumables like printer head lamp tubing, paper roll etc& Breakable parts are not covered by warranty and hence are chargeable, parts replaced due to negligence inopration will also be changed in every case			

Deals in: MEDICAL INSTRUMENTS,POWERSUPPLIES, TELE COMMUNICATION, MEDICALREAGENT &KITS