

TRU CAL METROLOGY SERVICES

2-B, Civil Lines, Yukti Business Centre
Near Old Session Court, Ambala City -134003, (HR.)

(**)+91-171-2441128, **!+91-9215301700

Email: info@trucal.in | website: www.trucal.in

CERTIFICATE OF CALIBRATION

Calibrated For Name & Address:-

M/S: Reliable Diagnostic Centre (Agarwal Nursing Home)

Salarpur Road, Kurukshetra, Haryana - 136118

10.710.757	Service request date:- 22/07/2022
Service request no.:- 2022/07/0576	Service request date ZZIOTIZOZZ
Date of receipt of Unit Under Calibration: -25/07/2022	Condition of Unit Under Calibration:-ok
	Next Suggested date of calibration: - 26/07/2023
	Certificate no.:- TCMS/13203.22
Date of Issue:- 26/07/2022	Certificate no 1CNIS/13203.22

DESCRIPTION & IDENTIFICATION OF INSTRUMENT

Nomenclature:- Thermo Hygrometer	Range:- (-50) to 70°C/10 to 99 % RH		
Make:- R-Tek	Least count:- 0.1°C/ 1% RH Accuracy:- ± 1°C/± 5% RH		
Model no.:- Not Mentioned			
Serial no.:- Not Mentioned	Acceptance Criteria: - ± 1°C/± 5% RH		
ID no.:- RDC/ANH/HYGRO/05	Type:- Digital		
Location:	Department Name:		
Cali Procedure - SOP/THER/005	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1		
Environment Conditions.:- Temperature:- 25	± 4°C, Calibration done at: -At Lab		
Relative Humidity:- 30 to 75 % RH			

DETAILS OF STANDARD USED

Nomenclature	Make/Model	ID/Serial No.	ULR/Certificate No.	Calibration Valid up to
Temperature & Humidity Indicator	Radix / Humirad 71	TRU/CAL/RHTI01 112003850	CC257521 000018137F	07/09/2022

DISCIPLINE: - THERMAL (Temperature Calibration)

RESULTS

SERIAL	UUC. READING	STD. READING	EXPANDED
NO.	IN (°C)	IN (°C)	UNCERTAINTY (±)
1.	10.7	10.5	
2.	25.6	25.3	0.7°C
3.	45.9	45.5	
	(%RH.)	(%RH)	5
1	22	21.4	3.0 % RH
2.	43	41.7	5.5 76 KH
3.	87	85.4	

Remarks:-

- 1. This certificate refers only to the particular item submitted for calibration.
- 2. This certificate shall not be reproduced, except in full, without the written permission of Tru Cal.
- 3. Results reported are valid at the time of & under the stated conditions of measurement.
- 4. Laboratory Standards are traceable to National Standards.
- 5. The reported expanded uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by coverage factor k=2, which is for a normal distribution corresponds to a coverage probability and level of confidence of approximately 95%.

Calibrated by:

Approved by:

Page 1 of 1 FC/CAL/051